



P.O. Box 9
Paragould, AR 72451

Ph: 870.239.7700
Fax: 870.239.7798
paragould.com

November 3, 2022

DMR-QA Officer
Arkansas Dept of Energy and Environment
5301 Northshore Drive
North Little Rock, AR 72118

Re: DMR-QA 42 Corrective Action Plan

Paragould Light Water & Cable participated in the DMR-QA Study 42 and our graded report was received on August 12, 2022. Unfortunately, (2) analyte results for Cadmium and Chromium were found to be unacceptable. PLWC is not required to test for these (2) analytes as they are not required by our NPDES Permit AR0033766, however; the PLWC lab conducted additional testing for these (2) analytes after implementing the corrective action which included:

1. Obtain new standard solutions for both Cadmium and Chromium
2. Purchased and installed new bulbs for our Agilent 240FS-AA metals analyzer for both Cadmium and Chromium

Tests were conducted and results were submitted to Phenova on September 30, 2022. The graded results of that test were received on October 3, 2022 with an evaluation of "Acceptable" for both Cadmium and Chromium.

If you have any questions please contact me at 239-7723 or dromine@paragould.com. Thank you.

Sincerely,

David Romine
Environmental Services Manager