


H₂S Assessment

GP - Crossett, AR			Constituent type: H ₂ S				CV _{ub} (%)	Bias (%)			
Date	Meas Val (Y)	Audit Val (X)	d (Eqn. 1)	25th Percentile	d ²	d	d ²				
12/16/2015 13:00	71.0	70.0	1.4	-0.286	2.041	1.429	2.041				
12/17/2015 13:00	69.8	70.0	-0.3	75th Percentile	0.082	0.286	0.082				
12/18/2015 13:00	69.8	70.0	-0.3	1.143	0.082	0.286	0.082				
12/19/2015 13:00	69.7	70.0	-0.4		0.184	0.429	0.184				
12/20/2015 13:00	69.9	70.0	-0.1		0.020	0.143	0.020				
12/21/2015 13:00	70.8	70.0	1.1		1.306	1.143	1.306				
12/22/2015 13:00	70.4	70.0	0.6		0.327	0.571	0.327				
12/23/2015 13:00	71.4	70.0	2.0		4.000	2.000	4.000				
12/24/2015 13:00	70.7	70.0	1.0		1.000	1.000	1.000				
12/25/2015 13:00	70.6	70.0	0.9		0.735	0.857	0.735				
12/26/2015 13:00	70.7	70.0	1.0		1.000	1.000	1.000				
12/27/2015 13:00	70.8	70.0	1.1		1.306	1.143	1.306				
12/29/2015 13:00	69.3	70.0	-1.0		1.000	1.000	1.000				
							n	S_d	S_{d2}	Σ d 	"AB" (Eqn 4)
							13	0.881	1.084	11.286	0.868
							n-1	Σd	Σd²	Σ d ²	"AS" (Eqn 5)
							12	7.000	13.082	13.082	0.523
									Bias (%) (Eqn 3)	Both Signs Positive	
									1.13	FALSE	
									Signed Bias (%)	Both Signs Negative	
									+/-1.13	FALSE	
									CV (%) (Eqn 2)		
									1.22		
									Upper Probability Limit	Lower Probability Limit	
									2.27	-1.19	



Meteorological Summary

