

November 15, 2016



6312 NW 18th Drive
Suite 100
Gainesville, FL 32653

352.378.0332 PHONE
352.378.0354 FAX

www.TRCsolutions.com

November 15, 2016

Ms. Lori Simmons
Arkansas Department of Health
4815 West Markham Street
Little Rock, Arkansas 72205
Via email Lori.Simmons@arkansas.gov

Re: Georgia-Pacific, Crossett Mill - Biweekly Air Monitoring Report for Hydrogen Sulfide

Dear Ms. Simmons,

Following is the biweekly data summary for the Georgia-Pacific (GP) hydrogen sulfide (H₂S) and meteorological monitoring program, at the GP Crossett mill, covering the calendar period of October 19th through November 1st.

Summary of Results

Included in this report are three plots presenting H₂S concentrations calculated with varied rolling average periods (30-minute, 8-hour, and 24-hour). Please note, observed H₂S concentrations were elevated on October 22nd and 30th. The highest 30-minute average concentration on the 22nd was 113.18 ppb and 90.56 ppb on the 30th.

Also included in this report is a summary of results from the daily 1-point QC checks performed during this biweekly period. The QAPP establishes goals for precision and bias as a coefficient of variation (CV) <10% and ± 10%, respectively. Precision and bias are calculated in accordance with 40 CFR Part 58 Appendix A, Section 4.1.

There were no occurrences of data loss, other than those resulting from automated daily 1-point QC and weekly calibration checks. Results for all automated daily 1-point QC checks fall within the acceptable range, indicating the H₂S monitor was operating in accordance with the QAPP.

Fourteen-day time series plots for all recorded meteorological (met) parameters are presented in the final table. All met parameters have 100% data capture for this report period.

Please feel free to contact me if you have any questions or need any additional data.

Sincerely,




November 15, 2016




Jonathan Bowser
Manager, Air Quality and Meteorological Monitoring

Air Measurements – Gainesville Office
6312 NW 18th Drive, Suite 100
Gainesville, Florida 32653
(352) 260-1162
Email: jbowser@trcsolutions.com


CC: Becky Keough, ADEQ Director via email: keogh@adeq.state.ar.us
Kara Allen, Environmental Engineer, USEPA Region 6 via email Allen.Kara@epa.gov



H2S 8 Hr Rolling Avg Georgia Pacific Crossett, AR



H₂S 24 Hr Rolling Avg Georgia Pacific Crossett, AR



H₂S Assessment

GP - Crossett, AR			Compound of Interest: H ₂ S					CV _{ub} (%)	Bias (%)	
Date	Meas Val (Y)	Audit Val (X)	d (Eqn. 1)	25th Percentile	d ²	d	d ²			
10/19/2016 13:00	70.4	70.0	0.6	-2.893	0.327	0.571	0.327			
10/20/2016 13:00	69.9	70.0	-0.1	75th Percentile	0.020	0.143	0.020	n	S _d	S _{d2}
10/21/2016 13:00	67.5	70.0	-3.6	-1.464	12.755	3.571	12.755		14	1.242
10/22/2016 13:00	67.4	70.0	-3.7		13.796	3.714	13.796	n-1	Σd	Σd ²
10/23/2016 13:00	68.4	70.0	-2.3		5.224	2.286	5.224	13	-28.429	77.776
10/24/2016 13:00	68.9	70.0	-1.6		2.469	1.571	2.469			
10/25/2016 13:00	67.6	70.0	-3.4		11.755	3.429	11.755			
10/26/2016 13:00	68.2	70.0	-2.6		6.612	2.571	6.612			
10/27/2016 13:00	67.9	70.0	-3.0		9.000	3.000	9.000	CV (%) (Eqn 2)		
10/28/2016 13:00	69.1	70.0	-1.3		1.653	1.286	1.653	1.69		
10/29/2016 13:00	69.0	70.0	-1.4		2.041	1.429	2.041			
10/30/2016 13:00	68.7	70.0	-1.9		3.449	1.857	3.449			
10/31/2016 13:00	68.7	70.0	-1.9		3.449	1.857	3.449			
11/1/2016 13:00	68.4	70.0	-2.3		5.224	2.286	5.224			
									</	

Meteorological Summary

