

6312 NW 18th Drive Suite 100 Gainesville, FL 32653

352.378.0332 PHONE 352.378.0354 FAX

www.TRCsolutions.com

June 23, 2017

Ms. Lori Simmons Arkansas Department of Health 4815 West Markham Street Little Rock, Arkansas 72205 Via email Lori.Simmons@arkansas.gov

Re: Georgia-Pacific, Crossett Mill - Biweekly Air Monitoring Report for Hydrogen Sulfide

Dear Ms. Simmons,

Following is the biweekly data summary for the Georgia-Pacific (GP) hydrogen sulfide (H₂S) and meteorological monitoring program, at the GP Crossett mill, covering the calendar period of May 31, 2017 through June 13, 2017.

Summary of Results

Included in this report are three plots presenting H₂S concentrations calculated with varied rolling average periods (30-minute, 8-hour, and 24-hour).

Also included in this report is a summary of results from the daily 1-point QC checks performed during this biweekly period. The QAPP establishes goals for precision and bias as a coefficient of variation (CV) <10% and \pm 10%, respectively. Precision and bias are calculated in accordance with 40 CFR Part 58 Appendix A, Section 4.1.

Additionally, weekly automated zero adjustments have been put in place beginning February 1, 2017, so as to limit the effect of the analyzer's zero drift. During this reporting period there were a total of eight zero checks performed; all within the acceptable range of \pm 1.5 ppb, as defined in the QAPP. Results for these zero checks are presented below.

Date	Zero Check	Date	Zero Check		
6/1/2017	-0.2	6/10/2017	0.3		
6/6/2017	0.0	6/11/2017	0.1		
6/8/2017	0.4	6/12/2017	0.1		
6/9/2017	0.3	6/13/2017	0.4		

There were multiple occurrences of data loss during this monitoring period, in addition to those



resulting from automated daily 1-point QC and weekly calibration checks. A multipoint calibration of the H_2S analyzer was performed June 6^{th} , resulting in the loss of approximately seven and a half hours of data loss. For an unknown reason daily automated 1-point QC checks have been low since June 7^{th} , therefore TRC performed manual multipoint checks (zero, \sim 70 ppb, and \sim 400 ppb) on a daily basis. The daily manual checks are responsible for approximately and hour and a half of data loss each day, from June 7^{th} – June 13^{th} . Results from the manual checks fall within the acceptable range, indicating the H_2S monitor was operating in accordance with the QAPP. These results were used in calculating the CV as shown in the table that follows. TRC has made plans to visit the site before the end of the month to identify and rectify the source of the automated calibration issue.

Fourteen-day time series plots for all recorded meteorological (met) parameters are presented in the final table. There was a single period of missing met data on June 6^{th} (less than one and a half hours) due to maintenance on the sensors.

Please feel free to contact me if you have any questions or need any additional data.

Sincerely,

Jonathan Bowser


Manager, Air Quality and Meteorological Monitoring

Air Measurements – Gainesville Office 6312 NW 18th Drive, Suite 100 Gainesville, Florida 32653 (352) 260-1162


Email: jbowser@trcsolutions.com

CC: Becky Keough, ADEQ Director via email: keogh@adeq.state.ar.us Kara Allen, Environmental Engineer, USEPA Region 6 via email Allen.Kara@epa.gov
















					H_2S	Asses	ssment	t					
GI	- Crossett, AF	<u> </u>	Compound	of Interest: H ₂ S					CV _{ub} (%)		Bias (%)		
Date	Meas Val (Y)	Audit Val (X)	d (Eqn. 1)	25th Percentile	d²	d	d ²						
5/31/2017 13:00	67.1	70.0	-4.1	-3.321	17.163		17.163						
6/1/2017 13:00	66.3	70.0	-5.3	75th Percentile	27.939	5.286	27.939	n	S _d	S _{d2}	∑ d	"AB" (Eqn 4)	
6/2/2017 13:00	67.5	70.0	-3.6	4.393	12.755	3.571	12.755	14	4.174	8.981	55.725	3.98	
6/3/2017 13:00	68.2	70.0	-2.6		6.612	2.571	6.612	n-1	∑d	$\sum d^2$	$\sum \mathbf{d} ^2$	"AS" (Eqn 5)	
6/4/2017 13:00	68.5	70.0	-2.1		4.592	2.143	4.592	13	12.846	238.287	238.287	1.12	
6/5/2017 13:00	72.4	70.0	3.4		11.755	3.429	11.755						
6/6/2017 13:00	49.1	51.0	-3.7		13.879	3.725	13.879				Bias (%) (Eqn 3)	Both Signs Positive	
6/7/2017 13:00	73.1	70.0	4.4		19.612	4.429	19.612				4.51	FALSE	
6/8/2017 13:00	71.7	70.0	2.4		5.898	2.429	5.898		CV (%) (Eqn 2)		Signed Bias (%)	Both Signs Negativ	
6/9/2017 13:00	72.6	70.0	3.7		13.796	3.714	13.796		5.67		+/-4.51	FALSE	
6/10/2017 13:00	73.0	70.0	4.3		18.367	4.286	18.367						
6/11/2017 13:00	73.3	70.0	4.7		22.224		22.224	_	Upper Probabil	ity Limit	Lower Probabilit	y Limit	
6/12/2017 13:00	73.9				31.041		31.041		9.1		-7.26	-	
6/13/2017 13:00	74.0	70.0	5.7		32.653	5.714	32.653						
							15.0		Perce	Percent Differences			
							5.0			*			
						0.0			_/				
						-5.0	•	-	-	¥			
						-10.0							
							-15.0 ¹						



