



July 23, 2018

Georgia Pacific Corporation
ATTN: Ms. Rachel Johnson
100 Supply Road
Drop Point 33
Crossett, AR 71635

Re: Analysis of Copper (Outfall 001)
American Interplex Corporation Control No. 224941

Dear Ms. Johnson,

One (1) water sample identified as Outfall 001, collected on July 3, 2018, was submitted to American Interplex Corporation on July 6, 2018, for analysis of Copper. The sample was prepared for analysis and analyzed by EPA 200.8 (ICP-MS) with a result of 50 ug/L. The sample was re-prepped, and re-analyzed for Copper by EPA 200.7 (ICP-AES) with a result of 12 ug/L. Additionally, the sample was re-prepped and re-analyzed for Copper by EPA 200.8 (ICP-MS), with a result of 12 ug/L. It is theorized that matrix interference and possible laboratory contamination caused the erroneous result of 50 ug/L from the original analysis which was reported to Georgia Pacific Corporation.

If I can be of any further assistance with this matter, please feel free to contact me.

Sincerely,

AMERICAN INTERPLEX CORPORATION


Steve Bradford
Laboratory Director

SB