

NON-COMPLIANCE REPORT

Arkansas Department of Environmental Quality
NPDES Enforcement Section
5301 Northshore Drive
North Little Rock, AR 72118

RE: NPDES Permit No: AR0021792 Discharge Number: AFIN 08-00034

Facility: City of Berryville

Address: 1000 West Cedarvale Rd.

City: Berryville State: AR Zip: 727616

Contact: Mayo Miller, Jacobs Project Manager Phone: 479-443-3292

Date of Non-Compliance	Parameter Exceeded	Quantity or Loading	Quality or Concentration	Permit Limits
04/03/18	Total Dissolved Solids (TDS)	10,094	460	Report

We feel this problem was due to:

TDS accuracy for the analysis was outside allowable control limits. This sample could not be reanalyzed due to expired holding time of sample. The known standard value was 306 mg/L, the recovered value was 340 mg/L. The percent recovery was 111.1% and the control limit for accuracy is 110%. Data value is being reported and part of the DMR calculation per Miles Johnson at ADEQ.

We plan on correcting the problem in this manner:

Reviewed the analytical SOP. QC data entry/ calculation is required at the completion of the analysis.

Time estimated that it will take to correct problem:

This issue was discovered on April 12, 2018 and corrected on April 12, 2018

Sincerely,

Jay Lee Admin. Assistant

Authorized Signature

5-14-18

Date

Certification Statement: I certify under penalty of law that this document and all attachments were prepared under my direction or supervision in accordance with a system designed to assure that qualified personnel properly gather and evaluate the information submitted. Based on my inquiry of the person or persons who manage the system, or those persons directly responsible for gathering the information, the information submitted is, to the best of my knowledge and belief, true, accurate, and complete. I am aware that there are significant penalties for submitting false information, including the possibility of fine and imprisonment for knowing violations. (Revised March 2016)