

268399

Environmental Enterprise Group, Inc.
220 North Knoxville
Russellville, Arkansas 72801
(479) 968-6767 Fax (479) 968-1956

4444 - 058366



Environmental Enterprise Group, Inc.
PROVIDING CUSTOMIZED SERVICES NATIONWIDE

Company Name: Clarksville Connected Utilities Phone #: PO 1807

Project Name or Number: CN Third Quarter Retest

Sampling Personnel Signature(s): Chris Cash Printed: Conner Cash

Sample I.D.	Date	Time	Comp	Grab	Cont. Type		# of Containers	Method Preserved		Sample Matrix				Requested Analysis	Laboratory Control Number	Remarks (Please note special detection limits below.)
					Plast.	Glass		HNO3	NaOH	HCL	Ice	None	Water			
Effluent	8-25-22	0625	X	X	X		1	X	X	X				X	0822199	
Equip New	8-25-22	0624	X	X			1	X	X	X				X	0822200	
Equip Blvd	8-25-22	0626	X	X			1	X	X	X				X	0822201	
Relinquished by:	<u>Conner Cash</u>		Date:	8-25-22	Time:	0100										
Received by:	<u>[Signature]</u>		Date:	8/25/22	Time:	600										
Relinquished by:	<u>[Signature]</u>		Date:	8/25/22	Time:	1100										
Comments:	<u>4, 1/2</u>															

Clarksville Connected Utilities
 Post Office Box 1807
 Clarksville, AR 72830

ANALYTICAL RESULTS

AIC No. 268399-1

Sample Identification: L444-058366 0822199 Effluent 25-Aug-2022 0625

<u>Analyte</u>	<u>Result</u>	<u>RL</u>	<u>Units</u>	<u>Qualifier</u>
Total Cyanide	< 0.01	0.01	mg/l	
SM 4500-CN C,E 2016	Prep: 29-Aug-2022 0923 by 376	Analyzed: 29-Aug-2022 1129 by 376	Batch: W80660	

AIC No. 268399-2

Sample Identification: L444-058366 0822200 Equip New 25-Aug-2022 0624

<u>Analyte</u>	<u>Result</u>	<u>RL</u>	<u>Units</u>	<u>Qualifier</u>
Total Cyanide	< 0.01	0.01	mg/l	
SM 4500-CN C,E 2016	Prep: 29-Aug-2022 0923 by 376	Analyzed: 29-Aug-2022 1135 by 376	Batch: W80660	

AIC No. 268399-3

Sample Identification: L444-058366 0822201 Equip Braid 25-Aug-2022 0626

<u>Analyte</u>	<u>Result</u>	<u>RL</u>	<u>Units</u>	<u>Qualifier</u>
Total Cyanide	< 0.01	0.01	mg/l	
SM 4500-CN C,E 2016	Prep: 29-Aug-2022 0923 by 376	Analyzed: 29-Aug-2022 1136 by 376	Batch: W80660	

August 29, 2022
 Control No. 268399
 Page 4 of 4

Clarksville Connected Utilities
 Post Office Box 1807
 Clarksville, AR 72830

LABORATORY CONTROL SAMPLE RESULTS

Analyte	Spike Amount	%	Limits	RPD	Limit	Batch	Preparation Date	Analysis Date	Dil	Qual
Total Cyanide	0.1 mg/l	89.6	81.9-118			W80660	29Aug22 0924 by 376	29Aug22 1128 by 376		

MATRIX SPIKE SAMPLE RESULTS

Analyte	Sample	Spike Amount	%	Limits	Batch	Preparation Date	Analysis Date	Dil	Qual
Total Cyanide	268399-1	0.1 mg/l	91.4	65.2-124	W80660	29Aug22 0924 by 376	29Aug22 1131 by 376		
	268399-1	0.1 mg/l	85.0	65.2-124	W80660	29Aug22 0924 by 376	29Aug22 1133 by 376		
Relative Percent Difference:			7.00	13.2	W80660				

LABORATORY BLANK RESULTS

Analyte	Result	RL	LOQ	QC Sample	Preparation Date	Analysis Date	Qual
Total Cyanide	< 0.0076 mg/l	0.0076	0.01	W80660-1	29Aug22 0924 by 376	29Aug22 1126 by 376	