

NON-COMPLIANCE REPORT

Arkansas Department of Environmental Quality
 NPDES Enforcement Section
 5301 Northshore Drive
 North Little Rock, AR 72118

RE: NPDES Permit No: AR0022187 Discharge Number: 001

Facility: Clarksville Light & Water

Address: 1305 South Crawford Street

City: Clarksville State: AR Zip: 72830

Contact: Gregg Rainey Phone: 479-754-7929

Date of Non-Compliance	Parameter Exceeded	Quantity or Loading	Quality or Concentration	Permit Limits <i>mo/avg</i>
12-26-2012	C-BOD	4.43 mg/L	28.67 mg/L	10

Everything in compliance glucose was too low.

We feel this problem was due to:

Incubator failure

We plan on correcting the problem in this manner:

HAD Incubator serviced

Time estimated that it will take to correct problem:

Fixed now

Sincerely,

Gregg Rainey
 Authorized Signature

12-27-2012
 Date