

CITY OF YELLVILLE  
WATER AND SEWER DEPARTMENT  
P.O. BOX 647  
YELLVILLE, AR 72687  
(870) 449-5210

January 21, 2022

Arkansas Department of Environmental Quality  
Enforcement Section/ Water Division  
5301 Northshore Drive  
North Little Rock, AR 72118

RE: Arkansas NPDES Permit Number AR 0034037

Dear Ms. Harbin,

This letter is to inform you of the City of Yellville's non-compliance with effluent limitations for the month of December 2021.

1. Description of Violations:

- a. Dissolved Oxygen, 6.0 (Inst. Min.), reported 5.2.

2. Reason for Violations:

- a. I believe ESC lab personnel were using a DO probe that required the user to slowly stir the probe in the sample being measured. They did not stir the probe in the sample resulting in a very low DO reading.

3. Duration of the Violation:


- a. This happened only one time this month.

4. Corrective Action Being Taken to Eliminate Violations:

- a. We will pay close attention to the instruments lab personnel are using.

Thank you for your time and considerations. If you should have any questions, please call my office at (870) 449-5210.

Sincerely,



Shawn Lane, Mayor  
City of Yellville