

Environmental Services Company, Inc.

Corporate Office
13715 West Markham
Little Rock, AR 72211
Tel. (501)221-2565 Fax (501)221-1341

Northwest Arkansas Branch
1107 Century Avenue
Springdale, AR 72762
Tel. (479)750-1170 Fax (479)750-1172

Control Number: 2205010357
Customer Name : FTN & ASSOCIATES
Customer Number : 1626
Report Date : 05/17/22

Sample Date : 04/27/22
Sample Time : 0930
Sample Type : GRAB WATER
Sample From : OUTFALL001 CLIN HGCD

Collected By: UNKNOWN
Delivery By : KEVIN SCHANKE
Work Order :
Purchase Order :


Laboratory Analysis

Analysis			Laboratory Analysis				Quality Assurance		
Date	Time	By	Parameter	Result	Notes	Quantity	Method	Precision % RPD	Accuracy % Recovery
05/09	1652	ARA	Mercury, low level	< 0.500 ng/L			EPA 1631E	0.60	96.5
05/09	1356	NTR	Cadmium	4.06 ug/L			EPA 200.8	1.13	84.3 *

* QA data shown is from a different sample or standard on the same date.

All equipment used is checked and/or calibrated daily. All NPDES testing is conducted in accordance with 40 CFR Part 136. A minimum of 10% spiked and duplicate samples is run on each parameter where applicable for Quality Assurance purposes. Quality Assurance Plan on file with Arkansas Department of Environmental Quality. Analysis time indicates the time of the start of the analytical batch in which the specific sample was included.

Signature


Environmental Services Co., Inc.

MS

