



5301 Northshore Drive
North Little Rock, AR 72118
Telephone: 501-682-0744

Client Report For: Exxon Oil Spill 2014 0714- 0723
Attention:
Client Address:

,

Report Date: March 19, 2014
LAB ID: AR14MAR13-01
Comment:

Approved By: _____

Date: March 19, 2014

Client: Special Samples **Client Sample ID:** WS-BKG-002
Lab ID: 2014-0714 **Collection Date:** 3/13/2014 10:53:00 AM
Matrix: Water

Analyses

<i>Dissolved Metals by EPA 200.8</i>		<i>EPA 200.8</i>	<i>Batch: 14031804 Run: 1</i>		
	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	169	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	0.66	0.5	0.2		ug/L
Barium	26.0	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	7.52	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.15	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	0.58	0.5	0.05		ug/L
Copper	1.90	0.5	0.2		ug/L
Iron	492	20	5.0		ug/L
Lead	0.41	0.3	0.02		ug/L
Magnesium	1.72	0.02	0.01		mg/L
Manganese	54.1	0.3	0.07		ug/L
Nickel	2.14	0.5	0.15		ug/L
Potassium	1.11	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	5.96	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	6.02	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	0.58	0.5	0.3		ug/L
Zinc	9.00	1	0.3		ug/L
Hardness	14.9	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 10:49AM				

Total Metals by EPA 200.8 *EPA 200.8* *Batch: 14031805 Run: 1*

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Aluminum	1880	20	20		ug/L
Antimony	<10	10	5		ug/L
Arsenic	<1	1	0.5		ug/L
Barium	31.9	10	2.0		ug/L
Beryllium	<0.5	0.5	0.1		ug/L
Boron	<25	25	5.0		ug/L
Cadmium	<1	1	0.3		ug/L
Calcium	3.60	0.04	0.04		mg/L
Chromium	1.49	1	0.3		ug/L
Cobalt	<1	1	0.5		ug/L
Copper	2.20	1	0.5		ug/L
Iron	1370	20	10.0		ug/L
Lead	1.39	1	0.1		ug/L
Magnesium	1.80	0.1	0.1		mg/L
Manganese	56.9	1	0.2		ug/L
Nickel	2.54	2.5	0.5		ug/L
Potassium	1.32	1	0.05		mg/L
Selenium	<2	2	0.5		ug/L
Silver	<5	5	1.0		ug/L
Sodium	6.26	0.04	0.02		mg/L
Thallium	<2.5	2.5	0.05		ug/L
Vanadium	2.52	2.5	1.0		ug/L
Zinc	11.1	3	2.0		ug/L

Dilution Factor 1

Analyzed By Robert Graddy

Analysis Date/Time Mar 17 2014 3:37PM

Prep By

Prep Date/Time

Client: Special Samples

Client Sample ID: WS-001

Lab ID: 2014-0715

Collection Date: 3/13/2014 12:40:00 PM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	172	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	0.55	0.5	0.2		ug/L
Barium	24.4	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	8.56	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.38	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	0.51	0.5	0.05		ug/L
Copper	1.71	0.5	0.2		ug/L
Iron	470	20	5.0		ug/L
Lead	0.44	0.3	0.02		ug/L
Magnesium	1.68	0.02	0.01		mg/L
Manganese	51.9	0.3	0.07		ug/L
Nickel	1.92	0.5	0.15		ug/L
Potassium	1.13	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	5.69	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	6.13	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	0.50	0.5	0.3		ug/L
Zinc	8.06	1	0.3		ug/L
Hardness	15.4	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 2:12PM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	1450	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	21.2	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.40	0.04	0.04	mg/L
Chromium	1.30	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	1.85	1	0.5	ug/L
Iron	1060	20	10.0	ug/L
Lead	1.02	1	0.1	ug/L
Magnesium	1.61	0.1	0.1	mg/L
Manganese	102	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.11	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	5.72	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	8.87	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 3:44PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-002

Lab ID: 2014-0716

Collection Date: 3/13/2014 10:19:00 AM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	42.2	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	11.7	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	14.9	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	2.90	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	0.97	0.5	0.2		ug/L
Iron	81.4	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	1.53	0.02	0.01		mg/L
Manganese	2.98	0.3	0.07		ug/L
Nickel	0.57	0.5	0.15		ug/L
Potassium	2.15	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	2.91	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	5.95	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	<1	1	0.3		ug/L
Hardness	13.5	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 11:23AM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

Limit

Aluminum	346	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	15.8	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.37	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	<1	1	0.5	ug/L
Iron	596	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	1.64	0.1	0.1	mg/L
Manganese	53.7	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.36	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	6.22	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	5.09	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 3:50PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-003

Lab ID: 2014-0717

Collection Date: 3/13/2014 9:40:00 AM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	47.7	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	13.6	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	15.1	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.25	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	1.76	0.5	0.2		ug/L
Iron	77.4	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	1.52	0.02	0.01		mg/L
Manganese	28.0	0.3	0.07		ug/L
Nickel	0.66	0.5	0.15		ug/L
Potassium	2.28	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	3.35	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	5.78	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	1.42	1	0.3		ug/L
Hardness	14.4	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 11:29AM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	340	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	17.8	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.39	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	6.76	1	0.5	ug/L
Iron	511	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	1.64	0.1	0.1	mg/L
Manganese	85.2	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.38	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	6.07	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	<3	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 3:57PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-005

Lab ID: 2014-0718

Collection Date: 3/13/2014 11:51:00 AM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	35.4	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	10.7	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	17.8	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.63	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	0.76	0.5	0.2		ug/L
Iron	73.7	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	1.61	0.02	0.01		mg/L
Manganese	5.33	0.3	0.07		ug/L
Nickel	0.63	0.5	0.15		ug/L
Potassium	2.35	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	2.31	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	6.77	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	<1	1	0.3		ug/L
Hardness	15.7	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 11:55AM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	331	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	14.1	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.80	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	<1	1	0.5	ug/L
Iron	454	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	1.71	0.1	0.1	mg/L
Manganese	55.7	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.50	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	6.89	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	<3	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 4:36PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-006

Lab ID: 2014-0719

Collection Date: 3/13/2014 1:16:00 PM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	84.2	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	0.53	0.5	0.2		ug/L
Barium	13.8	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	13.5	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.42	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	0.92	0.5	0.2		ug/L
Iron	136	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	1.55	0.02	0.01		mg/L
Manganese	49.8	0.3	0.07		ug/L
Nickel	0.73	0.5	0.15		ug/L
Potassium	2.05	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	3.44	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	5.83	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	2.59	1	0.3		ug/L
Hardness	14.9	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 12:02PM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	725	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	18.3	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.76	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	1.22	1	0.5	ug/L
Iron	675	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	1.69	0.1	0.1	mg/L
Manganese	83.9	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.26	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	6.17	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	5.78	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 4:42PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-007

Lab ID: 2014-0720

Collection Date: 3/13/2014 12:51:00 PM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	197	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	0.53	0.5	0.2		ug/L
Barium	17.2	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	8.62	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.20	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	1.25	0.5	0.2		ug/L
Iron	306	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	1.44	0.02	0.01		mg/L
Manganese	152	0.3	0.07		ug/L
Nickel	1.11	0.5	0.15		ug/L
Potassium	1.61	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	4.91	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	5.25	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	6.31	1	0.3		ug/L
Hardness	13.9	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 12:22PM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	2390	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	1.24	1	0.5	ug/L
Barium	31.1	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.65	0.04	0.04	mg/L
Chromium	2.01	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	2.36	1	0.5	ug/L
Iron	1880	20	10.0	ug/L
Lead	1.70	1	0.1	ug/L
Magnesium	1.70	0.1	0.1	mg/L
Manganese	204	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	1.99	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	5.5	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	3.04	2.5	1.0	ug/L
Zinc	20.9	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 4:49PM			
Prep By				
Prep Date/Time				

Client: Special Samples

Client Sample ID: WS-008

Lab ID: 2014-0721

Collection Date: 3/13/2014 11:19:00 AM

Matrix: Water

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Aluminum	82.8	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	14.6	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	11.6	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	3.18	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	0.7	0.5	0.2		ug/L
Iron	102	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	2.60	0.02	0.01		mg/L
Manganese	124	0.3	0.07		ug/L
Nickel	1.4	0.5	0.15		ug/L
Potassium	0.966	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	7.86	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	6.29	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	19.2	1	0.3		ug/L
Hardness	18.6	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 12:28PM				

Total Metals by EPA 200.8

EPA 200.8

Batch: 14031805 Run: 1

Result Reporting MDL Qual Unit

		Limit		
Aluminum	1420	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	<1	1	0.5	ug/L
Barium	20.2	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	3.46	0.04	0.04	mg/L
Chromium	1.25	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	1.17	1	0.5	ug/L
Iron	675	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	2.74	0.1	0.1	mg/L
Manganese	129	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	1.24	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	6.64	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	23.6	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 4:55PM			
Prep By				
Prep Date/Time				

Client: Special Samples	Client Sample ID: Metals Field Blank
Lab ID: 2014-0722	Collection Date: 3/13/2014 9:22:00 AM
Matrix: Water	

Analyses

Dissolved Metals by EPA 200.8

EPA 200.8

Batch: 14031804 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Aluminum	<20	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	<2	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	<5	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	<0.03	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	<0.5	0.5	0.2		ug/L
Iron	<20	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	<0.02	0.02	0.01		mg/L
Manganese	<0.3	0.3	0.07		ug/L
Nickel	<0.5	0.5	0.15		ug/L
Potassium	<0.02	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	<0.05	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	<0.02	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	<1	1	0.3		ug/L
Hardness	<1	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Mar 17 2014 12:35PM				

Client: Special Samples

Client Sample ID: WS-BKG-002

Lab ID: 2014-0714

Collection Date: 3/13/2014 10:53:00 AM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	42.0	40-110			%
Nitrobenzene-d5 (% Recovery)	81.7	50-110			%
2-Fluorobiphenyl (% Recovery)	82.3	50-110			%
2,4,6-Tribromophenol (% Recovery)	96.5	40-110			%
Terphenyl-d14 (% Recovery)	98.2	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Arkansas Department of Environmental Quality
 5301 Northshore Drive
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
 Ruehr@adeq.state.ar.us
 501-682-0955

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.815	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 1:20 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-001

Lab ID: 2014-0715

Collection Date: 3/13/2014 12:40:00 PM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	50.0	40-110			%
Nitrobenzene-d5 (% Recovery)	79.3	50-110			%
2-Fluorobiphenyl (% Recovery)	84.8	50-110			%
2,4,6-Tribromophenol (% Recovery)	90.3	40-110			%
Terphenyl-d14 (% Recovery)	81.5	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Arkansas Department of Environmental Quality
 5301 Northshore Drive
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
 Ruehr@adeq.state.ar.us
 501-682-0955

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.474	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 1:48 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-002

Lab ID: 2014-0716

Collection Date: 3/13/2014 10:19:00 AM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	45.0	40-110			%
Nitrobenzene-d5 (% Recovery)	80.1	50-110			%
2-Fluorobiphenyl (% Recovery)	73.2	50-110			%
2,4,6-Tribromophenol (% Recovery)	87.9	40-110			%
Terphenyl-d14 (% Recovery)	92.7	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Arkansas Department of Environmental Quality
 5301 Northshore Drive
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
 Ruehr@adeq.state.ar.us
 501-682-0955

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.450	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 2:16 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-003

Lab ID: 2014-0717

Collection Date: 3/13/2014 9:40:00 AM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	44.2	40-110			%
Nitrobenzene-d5 (% Recovery)	79.2	50-110			%
2-Fluorobiphenyl (% Recovery)	81.6	50-110			%
2,4,6-Tribromophenol (% Recovery)	93.1	40-110			%
Terphenyl-d14 (% Recovery)	93.1	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.558	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 2:45 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-005

Lab ID: 2014-0718

Collection Date: 3/13/2014 11:51:00 AM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	46.9	40-110			%
Nitrobenzene-d5 (% Recovery)	79.1	50-110			%
2-Fluorobiphenyl (% Recovery)	76.7	50-110			%
2,4,6-Tribromophenol (% Recovery)	91.6	40-110			%
Terphenyl-d14 (% Recovery)	86.9	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.322	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 4:38 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-006

Lab ID: 2014-0719

Collection Date: 3/13/2014 1:16:00 PM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	46.4	40-110			%
Nitrobenzene-d5 (% Recovery)	78.3	50-110			%
2-Fluorobiphenyl (% Recovery)	77.0	50-110			%
2,4,6-Tribromophenol (% Recovery)	91.1	40-110			%
Terphenyl-d14 (% Recovery)	90.6	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.396	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 5:07 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-007

Lab ID: 2014-0720

Collection Date: 3/13/2014 12:51:00 PM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	40.1	40-110			%
Nitrobenzene-d5 (% Recovery)	78.6	50-110			%
2-Fluorobiphenyl (% Recovery)	73.3	50-110			%
2,4,6-Tribromophenol (% Recovery)	88.6	40-110			%
Terphenyl-d14 (% Recovery)	83.5	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Arkansas Department of Environmental Quality
 5301 Northshore Drive
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
 Ruehr@adeq.state.ar.us
 501-682-0955

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	1.04	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 5:35 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Client: Special Samples

Client Sample ID: WS-008

Lab ID: 2014-0721

Collection Date: 3/13/2014 11:19:00 AM

Matrix: Water

Analyses

Oil and Grease

EPA1664

Batch: 14031806 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Semi-Volatiles by GC/MS

EPA 3510C/EPA 8270D

Batch: 14031708 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
2-Fluorophenol (% Recovery)	45.4	40-110			%
Nitrobenzene-d5 (% Recovery)	80.1	50-110			%
2-Fluorobiphenyl (% Recovery)	88.3	50-110			%
2,4,6-Tribromophenol (% Recovery)	101	40-110			%
Terphenyl-d14 (% Recovery)	92.2	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	0.326	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	3/17/2014 6:03 PM			
Prep By	Ed Harris			
Prep Date/Time	3/14/2014 08:00			

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-BKG-002

Lab ID: 2014-0714

Collection Date: 3/13/2014 10:53:00 AM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	21.7	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:01				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-001

Lab ID: 2014-0715

Collection Date: 3/13/2014 12:40:00 PM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	15.9	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 11:59				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-002

Lab ID: 2014-0716

Collection Date: 3/13/2014 10:19:00 AM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	9.06	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:06				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-003

Lab ID: 2014-0717

Collection Date: 3/13/2014 9:40:00 AM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	6.37	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:10				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-005

Lab ID: 2014-0718

Collection Date: 3/13/2014 11:51:00 AM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	7.27	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:18				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-006

Lab ID: 2014-0719

Collection Date: 3/13/2014 1:16:00 PM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	11.5	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:23				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-007

Lab ID: 2014-0720

Collection Date: 3/13/2014 12:51:00 PM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	27.1	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:27				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Client: Special Samples

Client Sample ID: WS-008

Lab ID: 2014-0721

Collection Date: 3/13/2014 11:19:00 AM

Matrix: Water

Analyses

Turbidity

EPA 180.1

Batch: 14031702 Run: 1

	<u>Result</u>	<u>Reporting Limit</u>	<u>MDL</u>	<u>Qual</u>	<u>Unit</u>
Turbidity	13.2	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:33				

Client: Special Samples

Client Sample ID: WS-BKG-002

Lab ID: 2014-0714

Collection Date: 3/13/2014 10:53:00 AM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Dibromofluoromethane (% Recovery)	107	70-130			%
1,2-Dichloroethane-d4 (% Recovery)	97.8	70-130			%
Toluene-d8 (% Recovery)	97.6	70-130			%
4-Bromofluorobenzene (% Recovery)	103	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 4:30 PM			

Client: Special Samples

Client Sample ID: WS-001

Lab ID: 2014-0715

Collection Date: 3/13/2014 12:40:00 PM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	103	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	93.8	70-130			%
<i>Toluene-d8 (% Recovery)</i>	94.7	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	99.1	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 4:55 PM			

Client: Special Samples

Client Sample ID: WS-002

Lab ID: 2014-0716

Collection Date: 3/13/2014 10:19:00 AM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	102	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	96.2	70-130			%
<i>Toluene-d8 (% Recovery)</i>	96.0	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	103	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 5:20 PM			

Client: Special Samples

Client Sample ID: WS-003

Lab ID: 2014-0717

Collection Date: 3/13/2014 9:40:00 AM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	103	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	96.1	70-130			%
<i>Toluene-d8 (% Recovery)</i>	98.1	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	110	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 7:26 PM			

Client: Special Samples

Client Sample ID: WS-005

Lab ID: 2014-0718

Collection Date: 3/13/2014 11:51:00 AM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Dibromofluoromethane (% Recovery)	104	70-130			%
1,2-Dichloroethane-d4 (% Recovery)	96.3	70-130			%
Toluene-d8 (% Recovery)	93.2	70-130			%
4-Bromofluorobenzene (% Recovery)	101	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 5:45 PM			

Client: Special Samples

Client Sample ID: WS-006

Lab ID: 2014-0719

Collection Date: 3/13/2014 1:16:00 PM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	106	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	96.9	70-130			%
<i>Toluene-d8 (% Recovery)</i>	97.7	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	102	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 6:10 PM			

Client: Special Samples

Client Sample ID: WS-007

Lab ID: 2014-0720

Collection Date: 3/13/2014 12:51:00 PM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	104	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	97.3	70-130			%
<i>Toluene-d8 (% Recovery)</i>	97.0	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	105	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 6:35 PM			

Client: Special Samples

Client Sample ID: WS-008

Lab ID: 2014-0721

Collection Date: 3/13/2014 11:19:00 AM

Matrix: Water

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
<i>Dibromofluoromethane (% Recovery)</i>	104	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	100	70-130			%
<i>Toluene-d8 (% Recovery)</i>	99.9	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	109	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 7:01 PM			

Client: Special Samples	Client Sample ID: Volatiles Field Blank
Lab ID: 2014-0723	Collection Date: 3/13/2014 2:10:00 PM
Matrix: Water	

Analyses

Volatile Organics by GCMS

EPA 8260C

Batch: 14031406 Run: 1

	Result	Reporting Limit	MDL	Qual	Unit
Dibromofluoromethane (% Recovery)	105	70-130			%
1,2-Dichloroethane-d4 (% Recovery)	96.8	70-130			%
Toluene-d8 (% Recovery)	96.4	70-130			%
4-Bromofluorobenzene (% Recovery)	100	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	3/13/2014 4:04 PM			

Analytical Quality Control Results Report

Batch: 14031406	VOA - water
WS-003	LIMS ID: 2014-0717

Volatiles - water DUP

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Dibromofluoromethane (% Recovery)	108 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	99.4 %			70 - 130	
Toluene-d8 (% Recovery)	97.9 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	106 %			70 - 130	
Dichlorodifluoromethane (RPD)	0 %				0 - 20
Chloromethane (RPD)	0 %				0 - 20
Vinyl chloride (RPD)	0 %				0 - 20
Bromomethane (RPD)	0 %				0 - 20
Chloroethane (RPD)	0 %				0 - 20
Trichlorofluoromethane (RPD)	0 %				0 - 20
1,1-Dichloroethene (RPD)	0 %				0 - 20
Acetone (RPD)	0 %				0 - 20
Methylene chloride (RPD)	7.2 %				0 - 20
Methyl tert-butyl ether (RPD)	0 %				0 - 20
trans-1,2-Dichloroethene (RPD)	0 %				0 - 20
1,1-Dichloroethane (RPD)	0 %				0 - 20
Methyl ethyl ketone (RPD)	0 %				0 - 20
cis-1,2-Dichloroethene (RPD)	0 %				0 - 20
2,2-Dichloropropane (RPD)	0 %				0 - 20
Bromochloromethane (RPD)	0 %				0 - 20
Chloroform (RPD)	0 %				0 - 20
1,1,1-Trichloroethane (RPD)	0 %				0 - 20
1,1-Dichloropropene (RPD)	0 %				0 - 20
Carbon tetrachloride (RPD)	0 %				0 - 20
Benzene (RPD)	0 %				0 - 20
1,2-Dichloroethane (RPD)	0 %				0 - 20
Trichloroethene (RPD)	0 %				0 - 20
1,2-Dichloropropane (RPD)	0 %				0 - 20
Dibromomethane (RPD)	0 %				0 - 20
Bromodichloromethane (RPD)	0 %				0 - 20
cis-1,3-Dichloropropene (RPD)	0 %				0 - 20

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Methyl isobutyl ketone (RPD)	0 %	0 - 20
Toluene (RPD)	0 %	0 - 20
trans-1,3-Dichloropropene (RPD)	0 %	0 - 20
1,1,2-Trichloroethane (RPD)	0 %	0 - 20
2-Hexanone (RPD)	0 %	0 - 20
Tetrachloroethene (RPD)	0 %	0 - 20
1,3-Dichloropropane (RPD)	0 %	0 - 20
Dibromochloromethane (RPD)	0 %	0 - 20
1,2-Dibromoethane (EDB) (RPD)	0 %	0 - 20
Chlorobenzene (RPD)	0 %	0 - 20
Ethylbenzene (RPD)	0 %	0 - 20
1,1,1,2-Tetrachloroethane (RPD)	0 %	0 - 20
m,p-Xylene (RPD)	0 %	0 - 20
o-Xylene (RPD)	0 %	0 - 20
Styrene (RPD)	0 %	0 - 20
Bromoform (RPD)	0 %	0 - 20
Isopropylbenzene (RPD)	0 %	0 - 20
1,1,2,2-Tetrachloroethane (RPD)	0 %	0 - 20
1,2,3-Trichloropropane (RPD)	0 %	0 - 20
n-Propylbenzene (RPD)	0 %	0 - 20
Bromobenzene (RPD)	0 %	0 - 20
1,3,5-Trimethylbenzene (RPD)	0 %	0 - 20
2-Chlorotoluene (RPD)	0 %	0 - 20
4-Chlorotoluene (RPD)	0 %	0 - 20
tert-Butylbenzene (RPD)	0 %	0 - 20
1,2,4-Trimethylbenzene (RPD)	0 %	0 - 20
sec-Butylbenzene (RPD)	0 %	0 - 20
p-Isopropyltoluene (RPD)	0 %	0 - 20
1,3-Dichlorobenzene (RPD)	0 %	0 - 20
1,4-Dichlorobenzene (RPD)	0 %	0 - 20
n-Butylbenzene (RPD)	0 %	0 - 20
1,2-Dichlorobenzene (RPD)	0 %	0 - 20
1,2-Dibromo-3-chloropropane (RPD)	0 %	0 - 20
1,2,4-Trichlorobenzene (RPD)	0 %	0 - 20
Hexachlorobutadiene (RPD)	0 %	0 - 20
Naphthalene (RPD)	0 %	0 - 20
1,2,3-Trichlorobenzene (RPD)	0 %	0 - 20
Dilution Factor	1	
Analyzed By	Jeff Ruehr	
Analysis Date/Time	3/13/2014 7:51 PM	

WS-003 **LIMS ID: 2014-0717**

Volatiles - water MS

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Dibromofluoromethane (% Recovery)	106 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	99.9 %			70 - 130	
Toluene-d8 (% Recovery)	99.8 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	110 %			70 - 130	
1,1-Dichloroethene (% Recovery)	106 %			70 - 130	
Benzene (% Recovery)	101 %			70 - 130	
Trichloroethene (% Recovery)	97.8 %			70 - 130	
Toluene (% Recovery)	91.8 %			70 - 130	
Chlorobenzene (% Recovery)	99.7 %			70 - 130	
Dilution Factor	1				
Analyzed By	Jeff Ruehr				
Analysis Date/Time	3/13/2014 8:16 PM				

WS-003 **LIMS ID: 2014-0717**

Volatiles - water MSD

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Dibromofluoromethane (% Recovery)	105 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	98.9 %			70 - 130	
Toluene-d8 (% Recovery)	94.4 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	98.5 %			70 - 130	
1,1-Dichloroethene (% Recovery)	104 %			70 - 130	
1,1-Dichloroethene (RPD)	1.6 %				0 - 20
Benzene (RPD)	3.1 %				0 - 20
Benzene (% Recovery)	97.5 %			70 - 130	
Trichloroethene (% Recovery)	96.2 %			70 - 130	
Trichloroethene (RPD)	1.6 %				0 - 20
Toluene (% Recovery)	85.5 %			70 - 130	
Toluene (RPD)	7.0 %				0 - 20
Chlorobenzene (RPD)	6.5 %				0 - 20
Chlorobenzene (% Recovery)	93.5 %			70 - 130	
Dilution Factor	1				
Analyzed By	Jeff Ruehr				
Analysis Date/Time	3/13/2014 8:42 PM				

LCS	LIMS ID: 14031406-LCS-01
------------	---------------------------------

Volatiles - water LCS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Dibromofluoromethane (% Recovery)	102 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	99.7 %			70 - 130	
Toluene-d8 (% Recovery)	101 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	110 %			70 - 130	
Dichlorodifluoromethane (% Recovery)	83.1 %			60 - 130	
Chloromethane (% Recovery)	91.9 %			60 - 130	
Vinyl chloride (% Recovery)	92.6 %			60 - 130	
Bromomethane (% Recovery)	99.2 %			60 - 130	
Chloroethane (% Recovery)	102 %			60 - 130	
Trichlorofluoromethane (% Recovery)	98.4 %			60 - 130	
1,1-Dichloroethene (% Recovery)	97.7 %			60 - 130	
Acetone (% Recovery)	109 %			60 - 130	
Methylene chloride (% Recovery)	220 %			60 - 130	
Methyl tert-butyl ether (% Recovery)	95.9 %			60 - 130	
trans-1,2-Dichloroethene (% Recovery)	100 %			60 - 130	
1,1-Dichloroethane (% Recovery)	95.9 %			60 - 130	
Methyl ethyl ketone (% Recovery)	96.5 %			60 - 130	
cis-1,2-Dichloroethene (% Recovery)	102 %			60 - 130	
2,2-Dichloropropane (% Recovery)	113 %			60 - 130	
Bromochloromethane (% Recovery)	99.2 %			60 - 130	
Chloroform (% Recovery)	98.0 %			60 - 130	
1,1,1-Trichloroethane (% Recovery)	100 %			60 - 130	
1,1-Dichloropropene (% Recovery)	100 %			60 - 130	
Carbon tetrachloride (% Recovery)	98.3 %			60 - 130	
Benzene (% Recovery)	96.2 %			60 - 130	
1,2-Dichloroethane (% Recovery)	99.7 %			60 - 130	
Trichloroethene (% Recovery)	99.9 %			60 - 130	
1,2-Dichloropropane (% Recovery)	96.6 %			60 - 130	
Dibromomethane (% Recovery)	96.8 %			60 - 130	
Bromodichloromethane (% Recovery)	96.8 %			60 - 130	
cis-1,3-Dichloropropene (% Recovery)	98.6 %			60 - 130	
Methyl isobutyl ketone (% Recovery)	95.6 %			60 - 130	
Toluene (% Recovery)	98.1 %			60 - 130	
trans-1,3-Dichloropropene (% Recovery)	98.4 %			60 - 130	
1,1,2-Trichloroethane (% Recovery)	94.9 %			60 - 130	
2-Hexanone (% Recovery)	98.6 %			60 - 130	

Tetrachloroethene (% Recovery)	99.9 %	60 - 130
1,3-Dichloropropane (% Recovery)	94.5 %	60 - 130
Dibromochloromethane (% Recovery)	95.1 %	60 - 130
1,2-Dibromoethane (EDB) (% Recovery)	93.9 %	60 - 130
Chlorobenzene (% Recovery)	99.7 %	60 - 130
Ethylbenzene (% Recovery)	97.8 %	60 - 130
1,1,1,2-Tetrachloroethane (% Recovery)	95.4 %	60 - 130
m,p-Xylene (% Recovery)	98.5 %	60 - 130
o-Xylene (% Recovery)	97.7 %	60 - 130
Styrene (% Recovery)	95.3 %	60 - 130
Bromoform (% Recovery)	97.1 %	60 - 130
Isopropylbenzene (% Recovery)	103 %	60 - 130
1,1,1,2,2-Tetrachloroethane (% Recovery)	101 %	60 - 130
1,2,3-Trichloropropane (% Recovery)	97.3 %	60 - 130
n-Propylbenzene (% Recovery)	103 %	60 - 130
Bromobenzene (% Recovery)	100 %	60 - 130
1,3,5-Trimethylbenzene (% Recovery)	106 %	60 - 130
2-Chlorotoluene (% Recovery)	106 %	60 - 130
4-Chlorotoluene (% Recovery)	103 %	60 - 130
tert-Butylbenzene (% Recovery)	103 %	60 - 130
1,2,4-Trimethylbenzene (% Recovery)	105 %	60 - 130
sec-Butylbenzene (% Recovery)	109 %	60 - 130
p-Isopropyltoluene (% Recovery)	105 %	60 - 130
1,3-Dichlorobenzene (% Recovery)	107 %	60 - 130
1,4-Dichlorobenzene (% Recovery)	104 %	60 - 130
n-Butylbenzene (% Recovery)	106 %	60 - 130
1,2-Dichlorobenzene (% Recovery)	103 %	60 - 130
1,2-Dibromo-3-chloropropane (% Recovery)	103 %	60 - 130
1,2,4-Trichlorobenzene (% Recovery)	105 %	60 - 130
Hexachlorobutadiene (% Recovery)	113 %	60 - 130
Naphthalene (% Recovery)	98.9 %	60 - 130
1,2,3-Trichlorobenzene (% Recovery)	102 %	60 - 130
Dilution Factor	1	
Analyzed By	Jeff Ruehr	
Analysis Date/Time	3/13/2014 14:24	

Analytical Quality Control Results Report

Batch: 14031708	Semi-VOA water (Prep)
WS-003	LIMS ID: 2014-0717

Semi Volatiles - water DUP

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	3/14/2014 08:00				
2-Fluorophenol (% Recovery)	42.8 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	75.3 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	82.4 %			40 - 110	
2,4,6-Tribromophenol (% Recovery)	87.4 %			40 - 110	
Terphenyl-d14 (% Recovery)	86.8 %			40 - 125	
Methyl Methanesulfonate (RPD)	0 %				0 - 40
Ethyl methanesulfonate (RPD)	0 %				0 - 40
Phenol (RPD)	0 %				0 - 40
Aniline (RPD)	4.4 %				0 - 40
Bis(2-chloroethyl)ether (RPD)	0 %				0 - 40
2-Chlorophenol (RPD)	200 %				0 - 40
1,3-Dichlorobenzene (RPD)	0 %				0 - 40
1,4-Dichlorobenzene (RPD)	0 %				0 - 40
Benzyl alcohol (RPD)	18.2 %				0 - 40
1,2-Dichlorobenzene (RPD)	0 %				0 - 40
2-Methylphenol (RPD)	0 %				0 - 40
Acetophenone (RPD)	49.4 %				0 - 40
4-Methylphenol (RPD)	0 %				0 - 40
N-Nitrosodi-n-propylamine (RPD)	0 %				0 - 40
Hexachloroethane (RPD)	0 %				0 - 40
Nitrobenzene (RPD)	0 %				0 - 40
N-Nitrosopiperidine (RPD)	0 %				0 - 40
Isophorone (RPD)	0 %				0 - 40
2-Nitrophenol (RPD)	0 %				0 - 40
2,4-Dimethylphenol (RPD)	0 %				0 - 40
Bis(2-chloroethoxy)methane (RPD)	0 %				0 - 40
2,4-Dichlorophenol (RPD)	0 %				0 - 40
1,2,4-Trichlorobenzene (RPD)	0 %				0 - 40
Naphthalene (RPD)	0.2 %				0 - 40

4-Chloroaniline (RPD)	0 %	0 - 40
2,6-Dichlorophenol (RPD)	0 %	0 - 40
Hexachlorobutadiene (RPD)	0 %	0 - 40
N-Nitrosodibutylamine (RPD)	0 %	0 - 40
4-Chloro-3-methylphenol (RPD)	0 %	0 - 40
2-Methylnaphthalene (RPD)	1.3 %	0 - 40
1,2,4,5-Tetrachlorobenzene (RPD)	0 %	0 - 40
Hexachlorocyclopentadiene (RPD)	0 %	0 - 40
2,4,6-Trichlorophenol (RPD)	0 %	0 - 40
2,4,5-Trichlorophenol (RPD)	0 %	0 - 40
2-Chloronaphthalene (RPD)	0 %	0 - 40
1-Chloronaphthalene (RPD)	0 %	0 - 40
2-Nitroaniline (RPD)	0 %	0 - 40
Dimethyl phthalate (RPD)	0 %	0 - 40
2,6-Dinitrotoluene (RPD)	0 %	0 - 40
Acenaphthylene (RPD)	0 %	0 - 40
3-Nitroaniline (RPD)	0 %	0 - 40
Acenaphthene (RPD)	0 %	0 - 40
2,4-Dinitrophenol (RPD)	0 %	0 - 40
Pentachlorobenzene (RPD)	0 %	0 - 40
4-Nitrophenol (RPD)	0 %	0 - 40
Dibenzofuran (RPD)	0 %	0 - 40
2,4-Dinitrotoluene (RPD)	0 %	0 - 40
2,3,4,6-Tetrachlorophenol (RPD)	0 %	0 - 40
Diethyl phthalate (RPD)	1.2 %	0 - 40
Fluorene (RPD)	0 %	0 - 40
4-Chlorophenyl phenyl ether (RPD)	0 %	0 - 40
4-Nitroaniline (RPD)	0 %	0 - 40
4,6-Dinitro-2-methylphenol (RPD)	0 %	0 - 40
Diphenylamine (RPD)	0 %	0 - 40
Azobenzene (RPD)	0 %	0 - 40
4-Bromophenyl phenyl ether (RPD)	0 %	0 - 40
Hexachlorobenzene (RPD)	0 %	0 - 40
Pentachlorophenol (RPD)	0 %	0 - 40
Pentachloronitrobenzene (RPD)	0 %	0 - 40
Pronamide (RPD)	0 %	0 - 40
Phenanthrene (RPD)	0 %	0 - 40
Anthracene (RPD)	0 %	0 - 40
Carbazole (RPD)	0 %	0 - 40
Di-n-butyl phthalate (RPD)	4.5 %	0 - 40

Fluoranthene (RPD)	0 %	0 - 40
Pyrene (RPD)	0 %	0 - 40
Dimethylaminoazobenzene (RPD)	0 %	0 - 40
Butyl benzyl phthalate (RPD)	0 %	0 - 40
Benzo (a) anthracene (RPD)	0 %	0 - 40
Chrysene (RPD)	0 %	0 - 40
Bis(2-ethylhexyl)phthalate (RPD)	20.3 %	0 - 40
Di-n-octyl phthalate (RPD)	9.1 %	0 - 40
Benzo (b) fluoranthene (RPD)	0 %	0 - 40
7,12-Dimethylbenz (a) anthracene (RPD)	0 %	0 - 40
Benzo (k) fluoranthene (RPD)	0 %	0 - 40
Benzo (a) pyrene (RPD)	0 %	0 - 40
3-Methylcholanthrene (RPD)	0 %	0 - 40
Indeno (1,2,3-cd) pyrene (RPD)	0 %	0 - 40
Dibenzo (a,h) anthracene (RPD)	0 %	0 - 40
Benzo (g,h,i) perylene (RPD)	0 %	0 - 40
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	3/17/2014 3:13 PM	

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

Semi Volatiles - water MS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	3/14/2014 08:00				
2-Fluorophenol (% Recovery)	47.0 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	85.6 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	80.7 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	95.7 %			40 - 125	
Terphenyl-d14 (% Recovery)	94.2 %			40 - 125	
Phenol (% Recovery)	32.2 %			25 - 125	
2-Chlorophenol (% Recovery)	65.9 %			25 - 125	
1,4-Dichlorobenzene (% Recovery)	72.2 %			25 - 125	
N-Nitrosodi-n-propylamine (% Recovery)	86.2 %			25 - 125	
1,2,4-Trichlorobenzene (% Recovery)	69.8 %			25 - 125	
4-Chloro-3-methylphenol (% Recovery)	76.1 %			25 - 125	
Acenaphthene (% Recovery)	74.5 %			25 - 125	

4-Nitrophenol (% Recovery)	49.7 %	25 - 125
2,4-Dinitrotoluene (% Recovery)	84.7 %	25 - 125
Pentachlorophenol (% Recovery)	100 %	25 - 125
Pyrene (% Recovery)	83.4 %	25 - 125
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	3/17/2014 3:41 PM	

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

Semi Volatiles - water MSD

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	3/14/2014 08:00				
2-Fluorophenol (% Recovery)	47.1 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	85.5 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	85.6 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	96.8 %			40 - 125	
Terphenyl-d14 (% Recovery)	89.6 %			40 - 125	
Phenol (% Recovery)	30.8 %			25 - 125	
Phenol (RPD)	4.4 %				0 - 40
2-Chlorophenol (% Recovery)	67.8 %			25 - 125	
2-Chlorophenol (RPD)	2.8 %				0 - 40
1,4-Dichlorobenzene (RPD)	7.1 %				0 - 40
1,4-Dichlorobenzene (% Recovery)	67.2 %			25 - 125	
N-Nitrosodi-n-propylamine (% Recovery)	88.9 %			25 - 125	
N-Nitrosodi-n-propylamine (RPD)	3.1 %				0 - 40
1,2,4-Trichlorobenzene (% Recovery)	67.9 %			25 - 125	
1,2,4-Trichlorobenzene (RPD)	2.9 %				0 - 40
4-Chloro-3-methylphenol (% Recovery)	73.3 %			25 - 125	
4-Chloro-3-methylphenol (RPD)	3.7 %				0 - 40
Acenaphthene (% Recovery)	74.8 %			25 - 125	
Acenaphthene (RPD)	0.5 %				0 - 40
4-Nitrophenol (% Recovery)	44.9 %			25 - 125	
4-Nitrophenol (RPD)	10.2 %				0 - 40
2,4-Dinitrotoluene (% Recovery)	89.9 %			25 - 125	
2,4-Dinitrotoluene (RPD)	6.0 %				0 - 40
Pentachlorophenol (% Recovery)	97.6 %			25 - 125	

Pentachlorophenol (RPD)	2.9 %	0 - 40
Pyrene (% Recovery)	87.8 %	25 - 125
Pyrene (RPD)	5.1 %	0 - 40
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	3/17/2014 4:10 PM	

MB	LIMS ID: 14031708-MB-01
-----------	--------------------------------

Semi Volatiles - water MB

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	3/14/2014 08:00				
2-Fluorophenol (% Recovery)	45.9 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	86.7 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	93.7 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	90.6 %			40 - 125	
Terphenyl-d14 (% Recovery)	99.6 %			40 - 125	
Methyl Methanesulfonate	<0.2 ug/L	0.2	0.2		
Ethyl methanesulfonate	<0.2 ug/L	0.2	0.2		
Phenol	<0.2 ug/L	0.2	0.2		
Aniline	<0.2 ug/L	0.2	0.2		
Bis(2-chloroethyl)ether	<0.2 ug/L	0.2	0.2		
2-Chlorophenol	<0.2 ug/L	0.2	0.2		
1,3-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
1,4-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
Benzyl alcohol	<0.16 ug/L	0.16	0.16		
1,2-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
2-Methylphenol	<0.1 ug/L	0.1	0.1		
Acetophenone	<0.1 ug/L	0.1	0.1		
4-Methylphenol	<0.1 ug/L	0.1	0.1		
N-Nitrosodi-n-propylamine	<0.2 ug/L	0.2	0.2		
Hexachloroethane	<0.2 ug/L	0.2	0.2		
Nitrobenzene	<0.2 ug/L	0.2	0.2		
N-Nitrosopiperidine	<0.2 ug/L	0.2	0.2		
Isophorone	<0.1 ug/L	0.1	0.1		
2-Nitrophenol	<0.3 ug/L	0.3	0.3		
2,4-Dimethylphenol	<0.1 ug/L	0.1	0.1		

Bis(2-chloroethoxy)methane	<0.2 ug/L	0.2	0.2
2,4-Dichlorophenol	<0.2 ug/L	0.2	0.2
1,2,4-Trichlorobenzene	<0.12 ug/L	0.12	0.12
Naphthalene	<0.08 ug/L	0.08	0.08
4-Chloroaniline	<0.1 ug/L	0.1	0.1
2,6-Dichlorophenol	<0.2 ug/L	0.2	0.2
Hexachlorobutadiene	<0.2 ug/L	0.2	0.2
N-Nitrosodibutylamine	<0.2 ug/L	0.2	0.2
4-Chloro-3-methylphenol	<0.16 ug/L	0.16	0.16
2-Methylnaphthalene	<0.12 ug/L	0.12	0.12
1,2,4,5-Tetrachlorobenzene	<0.1 ug/L	0.1	0.1
Hexachlorocyclopentadiene	<0.16 ug/L	0.16	0.16
2,4,6-Trichlorophenol	<0.2 ug/L	0.2	0.2
2,4,5-Trichlorophenol	<0.2 ug/L	0.2	0.2
2-Chloronaphthalene	<0.1 ug/L	0.1	0.1
1-Chloronaphthalene	<0.1 ug/L	0.1	0.1
2-Nitroaniline	<0.2 ug/L	0.2	0.2
Dimethyl phthalate	<0.2 ug/L	0.2	0.2
2,6-Dinitrotoluene	<0.2 ug/L	0.2	0.2
Acenaphthylene	<0.08 ug/L	0.08	0.08
3-Nitroaniline	<0.2 ug/L	0.2	0.2
Acenaphthene	<0.1 ug/L	0.1	0.1
2,4-Dinitrophenol	<4 ug/L	4	4
Pentachlorobenzene	<0.12 ug/L	0.12	0.12
4-Nitrophenol	<2 ug/L	2	2
Dibenzofuran	<0.1 ug/L	0.1	0.1
2,4-Dinitrotoluene	<0.2 ug/L	0.2	0.2
2,3,4,6-Tetrachlorophenol	<0.6 ug/L	0.6	0.6
Diethyl phthalate	<0.2 ug/L	0.2	0.2
Fluorene	<0.1 ug/L	0.1	0.1
4-Chlorophenyl phenyl ether	<0.1 ug/L	0.1	0.1
4-Nitroaniline	<0.2 ug/L	0.2	0.2
4,6-Dinitro-2-methylphenol	<6 ug/L	6	6
Diphenylamine	<0.1 ug/L	0.1	0.1
Azobenzene	<0.08 ug/L	0.08	0.08
4-Bromophenyl phenyl ether	<0.2 ug/L	0.2	0.2
Hexachlorobenzene	<0.16 ug/L	0.16	0.16
Pentachlorophenol	<1 ug/L	1	1
Pentachloronitrobenzene	<0.2 ug/L	0.2	0.2
Pronamide	<0.2 ug/L	0.2	0.2

Phenanthrene	<0.08 ug/L	0.08	0.08
Anthracene	<0.08 ug/L	0.08	0.08
Carbazole	<0.1 ug/L	0.1	0.1
Di-n-butyl phthalate	1.55 ug/L	0.2	0.2
Fluoranthene	<0.08 ug/L	0.08	0.08
Pyrene	<0.08 ug/L	0.08	0.08
Dimethylaminoazobenzene	<0.2 ug/L	0.2	0.2
Butyl benzyl phthalate	<0.3 ug/L	0.3	0.3
Benzo (a) anthracene	<0.2 ug/L	0.2	0.2
Chrysene	<0.1 ug/L	0.1	0.1
Bis(2-ethylhexyl)phthalate	<0.3 ug/L	0.3	0.3
Di-n-octyl phthalate	<0.3 ug/L	0.3	0.3
Benzo (b) fluoranthene	<0.16 ug/L	0.16	0.16
7,12-Dimethylbenz (a) anthracene	<0.2 ug/L	0.2	0.2
Benzo (k) fluoranthene	<0.16 ug/L	0.16	0.16
Benzo (a) pyrene	<0.16 ug/L	0.16	0.16
3-Methylcholanthrene	<0.2 ug/L	0.2	0.2
Indeno (1,2,3-cd) pyrene	<0.2 ug/L	0.2	0.2
Dibenzo (a,h) anthracene	<0.16 ug/L	0.16	0.16
Benzo (g,h,i) perylene	<0.16 ug/L	0.16	0.16
Dilution Factor	1		
Analyzed By	Ed Harris		
Analysis Date/Time	3/17/2014 12:23		

LCS	LIMS ID: 14031708-LCS-01
------------	---------------------------------

Semi Volatiles - water LCS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	3/14/2014 08:00				
2-Fluorophenol (% Recovery)	52.0 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	86.1 %			50 - 125	
2-Fluorobiphenyl (% Recovery)	80.5 %			50 - 125	
2,4,6-Tribromophenol (% Recovery)	105 %			40 - 125	
Terphenyl-d14 (% Recovery)	95.1 %			50 - 125	
Methyl Methanesulfonate (% Recovery)	76.7 %			50 - 150	
Ethyl methanesulfonate (% Recovery)	98.3 %			50 - 150	
Phenol (% Recovery)	41.4 %			50 - 150	

Aniline (% Recovery)	46.5 %	50 - 150
Bis(2-chloroethyl)ether (% Recovery)	102 %	50 - 150
2-Chlorophenol (% Recovery)	92.7 %	50 - 150
1,3-Dichlorobenzene (% Recovery)	59.8 %	50 - 150
1,4-Dichlorobenzene (% Recovery)	62.0 %	50 - 150
Benzyl alcohol (% Recovery)	89.2 %	50 - 150
1,2-Dichlorobenzene (% Recovery)	66.0 %	50 - 150
2-Methylphenol (% Recovery)	84.3 %	50 - 150
Acetophenone (% Recovery)	96.9 %	50 - 150
4-Methylphenol (% Recovery)	74.8 %	50 - 150
N-Nitrosodi-n-propylamine (% Recovery)	101 %	50 - 150
Hexachloroethane (% Recovery)	45.4 %	50 - 150
Nitrobenzene (% Recovery)	91.8 %	50 - 150
N-Nitrosopiperidine (% Recovery)	105 %	50 - 150
Isophorone (% Recovery)	106 %	50 - 150
2-Nitrophenol (% Recovery)	95.9 %	50 - 150
2,4-Dimethylphenol (% Recovery)	7.0 %	50 - 150
Bis(2-chloroethoxy)methane (% Recovery)	90.3 %	50 - 150
2,4-Dichlorophenol (% Recovery)	91.8 %	50 - 150
1,2,4-Trichlorobenzene (% Recovery)	67.0 %	50 - 150
Naphthalene (% Recovery)	79.6 %	50 - 150
4-Chloroaniline (% Recovery)	72.7 %	50 - 150
2,6-Dichlorophenol (% Recovery)	99.6 %	50 - 150
Hexachlorobutadiene (% Recovery)	47.7 %	50 - 150
N-Nitrosodibutylamine (% Recovery)	111 %	50 - 150
4-Chloro-3-methylphenol (% Recovery)	98.0 %	50 - 150
2-Methylnaphthalene (% Recovery)	78.0 %	50 - 150
1,2,4,5-Tetrachlorobenzene (% Recovery)	78.8 %	50 - 150
Hexachlorocyclopentadiene (% Recovery)	52.7 %	50 - 150
2,4,6-Trichlorophenol (% Recovery)	115 %	50 - 150
2,4,5-Trichlorophenol (% Recovery)	106 %	50 - 150
2-Chloronaphthalene (% Recovery)	91.6 %	50 - 150
1-Chloronaphthalene (% Recovery)	85.6 %	50 - 150
2-Nitroaniline (% Recovery)	120 %	50 - 150
Dimethyl phthalate (% Recovery)	109 %	50 - 150
2,6-Dinitrotoluene (% Recovery)	108 %	50 - 150
Acenaphthylene (% Recovery)	101 %	50 - 150
3-Nitroaniline (% Recovery)	109 %	50 - 150
Acenaphthene (% Recovery)	90.6 %	50 - 150
2,4-Dinitrophenol (% Recovery)	125 %	50 - 150

Pentachlorobenzene (% Recovery)	84.4 %	50 - 150
4-Nitrophenol (% Recovery)	48.9 %	50 - 150
Dibenzofuran (% Recovery)	101 %	50 - 150
2,4-Dinitrotoluene (% Recovery)	116 %	50 - 150
2,3,4,6-Tetrachlorophenol (% Recovery)	116 %	50 - 150
Diethyl phthalate (% Recovery)	105 %	50 - 150
Fluorene (% Recovery)	104 %	50 - 150
4-Chlorophenyl phenyl ether (% Recovery)	99.6 %	50 - 150
4-Nitroaniline (% Recovery)	110 %	50 - 150
4,6-Dinitro-2-methylphenol (% Recovery)	112 %	50 - 150
Diphenylamine (% Recovery)	96.1 %	50 - 150
Azobenzene (% Recovery)	100 %	50 - 150
4-Bromophenyl phenyl ether (% Recovery)	91.9 %	50 - 150
Hexachlorobenzene (% Recovery)	84.6 %	50 - 150
Pentachlorophenol (% Recovery)	114 %	50 - 150
Pentachloronitrobenzene (% Recovery)	108 %	50 - 150
Pronamide (% Recovery)	110 %	50 - 150
Phenanthrene (% Recovery)	98.2 %	50 - 150
Anthracene (% Recovery)	97.8 %	50 - 150
Carbazole (% Recovery)	113 %	50 - 150
Di-n-butyl phthalate (% Recovery)	180 %	50 - 150
Fluoranthene (% Recovery)	112 %	50 - 150
Pyrene (% Recovery)	97.7 %	50 - 150
Dimethylaminoazobenzene (% Recovery)	142 %	50 - 150
Butyl benzyl phthalate (% Recovery)	129 %	50 - 150
Benzo (a) anthracene (% Recovery)	112 %	50 - 150
Chrysene (% Recovery)	106 %	50 - 150
Bis(2-ethylhexyl)phthalate (% Recovery)	136 %	50 - 150
Di-n-octyl phthalate (% Recovery)	152 %	50 - 150
Benzo (b) fluoranthene (% Recovery)	114 %	50 - 150
7,12-Dimethylbenz (a) anthracene (% Recovery)	93.5 %	50 - 150
Benzo (k) fluoranthene (% Recovery)	103 %	50 - 150
Benzo (a) pyrene (% Recovery)	107 %	50 - 150
3-Methylcholanthrene (% Recovery)	115 %	50 - 150
Indeno (1,2,3-cd) pyrene (% Recovery)	86.4 %	50 - 150
Dibenzo (a,h) anthracene (% Recovery)	84.5 %	50 - 150
Benzo (g,h,i) perylene (% Recovery)	80.7 %	50 - 150
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	3/17/2014 12:52	

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Analytical Quality Control Results Report

Batch: 14031805	ICP Metals - water (total)
WS-003	LIMS ID: 2014-0717

ICP Metals - water (Total) DUP

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum	394 ug/L	20	20		
Aluminum (RPD)	14.6 %				0 - 20
Antimony (RPD)	0 %				0 - 20
Antimony	<10 ug/L	5	10		
Arsenic	<1 ug/L	0.5	1		
Arsenic (RPD)	2.6 %				0 - 20
Barium (RPD)	3.6 %				0 - 20
Barium	18.4 ug/L	2	10		
Beryllium	<0.5 ug/L	0.1	0.5		
Beryllium (RPD)	14.3 %				0 - 20
Boron (RPD)	2.0 %				0 - 20
Boron	<25 ug/L	5	25		
Cadmium (RPD)	0 %				0 - 20
Cadmium	<1 ug/L	0.3	1		
Calcium	3.44 mg/L	0.04	0.04		
Calcium (RPD)	1.3 %				0 - 20
Chromium (RPD)	9.2 %				0 - 20
Chromium	<1 ug/L	0.3	1		
Cobalt	<1 ug/L	0.5	1		
Cobalt (RPD)	23.5 %				0 - 20
Copper (RPD)	146 %				0 - 20
Copper	1.06 ug/L	0.5	1		
Iron	587 ug/L	10	20		
Iron (RPD)	13.8 %				0 - 20
Lead (RPD)	13.5 %				0 - 20
Lead	<1 ug/L	0.1	1		
Magnesium	1.66 mg/L	0.1	0.1		
Magnesium (RPD)	1.5 %				0 - 20
Manganese	97 ug/L	0.2	1		
Manganese (RPD)	13 %				0 - 20
Nickel (RPD)	8.4 %				0 - 20
Nickel	<2.5 ug/L	0.5	2.5		
Potassium	2.41 mg/L	0.05	1		

Potassium (RPD)	1.3 %			0 - 20
Selenium (RPD)	21.8 %			0 - 20
Selenium	<2 ug/L	0.5	2	
Silver	<5 ug/L	1	5	
Silver (RPD)	0 %			0 - 20
Sodium (RPD)	0.2 %			0 - 20
Sodium	6.08 mg/L	0.02	0.04	
Thallium	<2.5 ug/L	0.05	2.5	
Thallium (RPD)	0 %			0 - 20
Vanadium (RPD)	34.0 %			0 - 20
Vanadium	<2.5 ug/L	1	2.5	
Zinc	<3 ug/L	2	3	
Zinc (RPD)	6.7 %			0 - 20
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 4:03PM			

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

ICP Metals - water (Total) MS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum (% Recovery)	93.3 %			70 - 130	
Antimony (% Recovery)	86.9 %			70 - 130	
Arsenic (% Recovery)	92.9 %			70 - 130	
Barium (% Recovery)	87.0 %			70 - 130	
Beryllium (% Recovery)	91.6 %			70 - 130	
Boron (% Recovery)	97.7 %			70 - 130	
Cadmium (% Recovery)	91.1 %			70 - 130	
Calcium (% Recovery)	89.0 %			70 - 130	
Chromium (% Recovery)	83.0 %			70 - 130	
Cobalt (% Recovery)	84.7 %			70 - 130	
Copper (% Recovery)	75.2 %			70 - 130	
Iron (% Recovery)	86.6 %			70 - 130	
Lead (% Recovery)	85.9 %			70 - 130	
Magnesium (% Recovery)	87.1 %			70 - 130	
Manganese (% Recovery)	100 %			70 - 130	
Nickel (% Recovery)	85 %			70 - 130	
Potassium (% Recovery)	89.2 %			70 - 130	
Selenium (% Recovery)	98.4 %			70 - 130	
Silver (% Recovery)	83.0 %			70 - 130	

Sodium (% Recovery)	86.3 %	70 - 130
Thallium (% Recovery)	84.4 %	70 - 130
Vanadium (% Recovery)	83.2 %	70 - 130
Zinc (% Recovery)	92.2 %	70 - 130
Dilution Factor	1	
Analyzed By	Robert Graddy	
Analysis Date/Time	Mar 17 2014 4:10PM	

WS-003 **LIMS ID: 2014-0717**

ICP Metals - water (Total) MSD

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum (% Recovery)	90.6 %			70 - 130	
Aluminum (RPD)	1.6 %				0 - 20
Antimony (% Recovery)	85.8 %			70 - 130	
Antimony (RPD)	1.3 %				0 - 20
Arsenic (% Recovery)	86.7 %			70 - 130	
Arsenic (RPD)	6.7 %				0 - 20
Barium (% Recovery)	86.8 %			70 - 130	
Barium (RPD)	0.2 %				0 - 20
Beryllium (% Recovery)	91.3 %			70 - 130	
Beryllium (RPD)	0.3 %				0 - 20
Boron (% Recovery)	98.0 %			70 - 130	
Boron (RPD)	0.3 %				0 - 20
Cadmium (% Recovery)	92.4 %			70 - 130	
Cadmium (RPD)	1.4 %				0 - 20
Calcium (% Recovery)	89.7 %			70 - 130	
Calcium (RPD)	0.6 %				0 - 20
Chromium (% Recovery)	84.3 %			70 - 130	
Chromium (RPD)	1.6 %				0 - 20
Cobalt (% Recovery)	84.8 %			70 - 130	
Cobalt (RPD)	0.1 %				0 - 20
Copper (% Recovery)	74.5 %			70 - 130	
Copper (RPD)	0.8 %				0 - 20
Iron (% Recovery)	91.4 %			70 - 130	
Iron (RPD)	1.4 %				0 - 20
Lead (% Recovery)	87.3 %			70 - 130	
Lead (RPD)	1.4 %				0 - 20
Magnesium (% Recovery)	87.7 %			70 - 130	
Magnesium (RPD)	0.6 %				0 - 20

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Manganese (% Recovery)	100 %	70 - 130	
Manganese (RPD)	0.3 %		0 - 20
Nickel (% Recovery)	85 %	70 - 130	
Nickel (RPD)	0.2 %		0 - 20
Potassium (% Recovery)	89.1 %	70 - 130	
Potassium (RPD)	0.1 %		0 - 20
Selenium (% Recovery)	99.3 %	70 - 130	
Selenium (RPD)	0.9 %		0 - 20
Silver (% Recovery)	83.1 %	70 - 130	
Silver (RPD)	0.1 %		0 - 20
Sodium (% Recovery)	85.7 %	70 - 130	
Sodium (RPD)	0.4 %		0 - 20
Thallium (% Recovery)	86.7 %	70 - 130	
Thallium (RPD)	2.6 %		0 - 20
Vanadium (% Recovery)	84.8 %	70 - 130	
Vanadium (RPD)	1.8 %		0 - 20
Zinc (% Recovery)	90.5 %	70 - 130	
Zinc (RPD)	1.7 %		0 - 20
Dilution Factor	1		
Analysis Date/Time	Mar 17 2014 4:16PM		
Analyzed By	Robert Graddy		

Analytical Quality Control Results Report

Batch: 14031804	ICP Metals - water (Diss.)
WS-003	LIMS ID: 2014-0717

ICP Metals - water (Dissolved) DUP

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Aluminum	30.4 ug/L	20	20		
Aluminum (RPD)	44.2 %				0 - 20
Antimony (RPD)	0 %				0 - 20
Antimony	<5 ug/L	1	5		
Arsenic	<0.5 ug/L	0.2	0.5		
Arsenic (RPD)	5.4 %				0 - 20
Barium (RPD)	14.9 %				0 - 20
Barium	15.7 ug/L	0.4	2		
Beryllium	<0.1 ug/L	0.04	0.1		
Beryllium (RPD)	0 %				0 - 20
Boron (RPD)	0.7 %				0 - 20
Boron	15.2 ug/L	2	5		
Cadmium (RPD)	0 %				0 - 20
Cadmium	<0.1 ug/L	0.05	0.1		
Calcium	3.30 mg/L	0.03	0.03		
Calcium (RPD)	1.5 %				0 - 20
Chromium (RPD)	11.1 %				0 - 20
Chromium	<0.5 ug/L	0.05	0.5		
Cobalt	<0.5 ug/L	0.05	0.5		
Cobalt (RPD)	0 %				0 - 20
Copper (RPD)	1.0 %				0 - 20
Copper	1.78 ug/L	0.2	0.5		
Iron	62.0 ug/L	5	20		
Iron (RPD)	22.1 %				0 - 20
Lead (RPD)	14.5 %				0 - 20
Lead	<0.3 ug/L	0.02	0.3		
Magnesium	1.52 mg/L	0.01	0.02		
Magnesium (RPD)	0.3 %				0 - 20
Manganese (RPD)	1.8 %				0 - 20
Manganese	27 ug/L	0.07	0.3		
Nickel	0.76 ug/L	0.15	0.5		
Nickel (RPD)	14 %				0 - 20
Potassium (RPD)	0.7 %				0 - 20

Potassium	2.30 mg/L	0.01	0.02	
Selenium	<1 ug/L	0.2	1	
Selenium (RPD)	0 %			0 - 20
Silicon Dioxide	3.30 mg/L	0.01	0.05	
Silicon Dioxide (RPD)	1.6 %			0 - 20
Silver (RPD)	0 %			0 - 20
Silver	<0.5 ug/L	0.02	0.5	
Sodium	5.79 mg/L	0.01	0.02	
Sodium (RPD)	0.2 %			0 - 20
Thallium (RPD)	0 %			0 - 20
Thallium	<0.5 ug/L	0	0.5	
Vanadium	<0.5 ug/L	0.3	0.5	
Vanadium (RPD)	12.3 %			0 - 20
Zinc (RPD)	94.0 %			0 - 20
Zinc	3.92 ug/L	0.3	1	
Hardness	14 mg/L	1	1	
Hardness (RPD)	0 %			0 - 20
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Mar 17 2014 11:36AM			

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

ICP Metals - water (Dissolved) MS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum (% Recovery)	84.8 %			70 - 130	
Antimony (% Recovery)	91.1 %			70 - 130	
Arsenic (% Recovery)	99.0 %			70 - 130	
Barium (% Recovery)	88.1 %			70 - 130	
Beryllium (% Recovery)	92.0 %			70 - 130	
Boron (% Recovery)	93.5 %			70 - 130	
Cadmium (% Recovery)	91.2 %			70 - 130	
Calcium (% Recovery)	87.5 %			70 - 130	
Chromium (% Recovery)	84.2 %			70 - 130	
Cobalt (% Recovery)	85.6 %			70 - 130	
Copper (% Recovery)	87.2 %			70 - 130	
Iron (% Recovery)	83.0 %			70 - 130	
Lead (% Recovery)	86.7 %			70 - 130	
Magnesium (% Recovery)	87.0 %			70 - 130	
Manganese (% Recovery)	81 %			70 - 130	

Nickel (% Recovery)	85 %	70 - 130
Potassium (% Recovery)	88.3 %	70 - 130
Selenium (% Recovery)	102 %	70 - 130
Silver (% Recovery)	84.4 %	70 - 130
Sodium (% Recovery)	86.3 %	70 - 130
Thallium (% Recovery)	84.4 %	70 - 130
Vanadium (% Recovery)	83.5 %	70 - 130
Zinc (% Recovery)	94.5 %	70 - 130
Dilution Factor	1	
Analyzed By	Robert Graddy	
Analysis Date/Time	Mar 17 2014 11:42AM	

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

ICP Metals - water (Dissolved) MSD

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum (% Recovery)	84.9 %			70 - 130	
Aluminum (RPD)	0.1 %				0 - 20
Antimony (% Recovery)	97.7 %			70 - 130	
Antimony (RPD)	7.1 %				0 - 20
Arsenic (% Recovery)	96.3 %			70 - 130	
Arsenic (RPD)	2.7 %				0 - 20
Barium (% Recovery)	88.3 %			70 - 130	
Barium (RPD)	0.2 %				0 - 20
Beryllium (% Recovery)	91.2 %			70 - 130	
Beryllium (RPD)	0.9 %				0 - 20
Boron (% Recovery)	94.1 %			70 - 130	
Boron (RPD)	0.6 %				0 - 20
Cadmium (% Recovery)	91.4 %			70 - 130	
Cadmium (RPD)	0.1 %				0 - 20
Calcium (% Recovery)	88.9 %			70 - 130	
Calcium (RPD)	1.2 %				0 - 20
Chromium (% Recovery)	83.6 %			70 - 130	
Chromium (RPD)	0.7 %				0 - 20
Cobalt (% Recovery)	85.5 %			70 - 130	
Cobalt (RPD)	0.2 %				0 - 20
Copper (% Recovery)	89.5 %			70 - 130	
Copper (RPD)	2.5 %				0 - 20
Iron (% Recovery)	82.1 %			70 - 130	
Iron (RPD)	0.8 %				0 - 20

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

Lead (% Recovery)	86.2 %	70 - 130	
Lead (RPD)	0.6 %		0 - 20
Magnesium (% Recovery)	87.8 %	70 - 130	
Magnesium (RPD)	0.8 %		0 - 20
Manganese (% Recovery)	80 %	70 - 130	
Manganese (RPD)	0.4 %		0 - 20
Nickel (% Recovery)	87 %	70 - 130	
Nickel (RPD)	2.1 %		0 - 20
Potassium (% Recovery)	87.9 %	70 - 130	
Potassium (RPD)	0.4 %		0 - 20
Selenium (% Recovery)	99.5 %	70 - 130	
Selenium (RPD)	2.2 %		0 - 20
Silver (% Recovery)	83.6 %	70 - 130	
Silver (RPD)	1.0 %		0 - 20
Sodium (% Recovery)	90.6 %	70 - 130	
Sodium (RPD)	2.9 %		0 - 20
Thallium (% Recovery)	81.8 %	70 - 130	
Thallium (RPD)	3.1 %		0 - 20
Vanadium (% Recovery)	82.8 %	70 - 130	
Vanadium (RPD)	0.9 %		0 - 20
Zinc (% Recovery)	96.0 %	70 - 130	
Zinc (RPD)	1.5 %		0 - 20
Dilution Factor	1		
Analyzed By	Robert Graddy		
Analysis Date/Time	Mar 17 2014 11:49AM		

Analytical Quality Control Results Report

Batch: 14031806	Oil and Grease - water
WS-003	LIMS ID: 2014-0717

Oil and Grease - water DUP

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (RPD)	0 %				0 - 20
Oil and Grease	<2.5 mg/L	2.5	2.5		
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

Oil and Grease - water MS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (% Recovery)	106 %			70 - 130	
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

WS-003	LIMS ID: 2014-0717
---------------	---------------------------

Oil and Grease - water MSD

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (% Recovery)	102 %			70 - 130	
Oil and Grease (RPD)	3.6 %				0 - 20
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

MB	LIMS ID: 14031806-MB-01
-----------	--------------------------------

Oil and Grease - water MB

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease	<2.5 mg/L	2.5	2.5		
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Arkansas Department of Environmental Quality
5301 Northshore Drive
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr
Ruehr@adeq.state.ar.us
501-682-0955

LCS

LIMS ID: 14031806-LCS-01

Oil and Grease - water LCS

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (% Recovery)	95.7 %			70 - 130	
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	03/17/2014 1600				

Analytical Quality Control Results Report

Batch: 14031702	Turbidity - water
WS-003	LIMS ID: 2014-0717

Turbidity - water DUP

Run: 1

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Turbidity	6.39 NTU	0.02	0.02		
Turbidity (RPD)	0.3 %				0 - 20
Dilution Factor	1				
Analyzed By	Penny Semberski				
Analysis Date/Time	3/14/2014 12:14				