



5301 Northshore Drive  
North Little Rock, AR 72118  
Telephone: 501-682-0744

**Client Report For:** Exxon Oil Spill 2015 1648-1652  
**Attention:**  
**Client Address:**

,

**Report Date:** August 19, 2015  
**LAB ID:** AR15JUN25-01  
**Comment:**

Approved By: \_\_\_\_\_

Date: August 19, 2015

**Client:** Special Samples

**Client Sample ID:** WS-001

**Lab ID:** 2015-1648

**Collection Date:** 6/25/2015 8:09:00 AM

**Matrix:** Water

**Analyses**

*Dissolved Metals by EPA 200.8*

*EPA 200.8*

*Batch: 15080703 Run: 1*

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Aluminum	<20	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	1.05	0.5	0.2		ug/L
Barium	8.47	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	13.9	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	5.05	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	0.68	0.5	0.2		ug/L
Iron	28.3	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	2.28	0.02	0.01		mg/L
Manganese	18.6	0.3	0.07		ug/L
Nickel	<0.5	0.5	0.15		ug/L
Potassium	2.26	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	2.94	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	2.90	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	<1	1	0.3		ug/L
Hardness	22.0	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Aug 7 2015 3:51AM				

*Total Metals by EPA 200.8*

*EPA 200.8*

*Batch: 15081303 Run: 1*

	<b><u>Result</u></b>	<b><u>Reporting Limit</u></b>	<b><u>MDL</u></b>	<b><u>Qual</u></b>	<b><u>Unit</u></b>
Aluminum	<20	20	20		ug/L
Antimony	<10	10	5		ug/L
Arsenic	1.70	1	0.5		ug/L
Barium	19.1	10	2.0		ug/L
Beryllium	<0.5	0.5	0.1		ug/L
Boron	<25	25	5.0		ug/L
Cadmium	<1	1	0.3		ug/L
Calcium	5.19	0.04	0.04		mg/L
Chromium	<1	1	0.3		ug/L
Cobalt	<1	1	0.5		ug/L
Copper	<1	1	0.5		ug/L
Iron	414	20	10.0		ug/L
Lead	<1	1	0.1		ug/L
Magnesium	2.32	0.1	0.1		mg/L
Manganese	381	1	0.2		ug/L
Nickel	<2.5	2.5	0.5		ug/L
Potassium	2.31	1	0.05		mg/L
Selenium	<2	2	0.5		ug/L
Silver	<5	5	1.0		ug/L
Sodium	3.55	0.04	0.02		mg/L
Thallium	<2.5	2.5	0.05		ug/L
Vanadium	<2.5	2.5	1.0		ug/L
Zinc	<3	3	2.0		ug/L

Dilution Factor 1

Analyzed By Robert Graddy

Analysis Date/Time Aug 11 2015 12:46AM

Prep By

Prep Date/Time

**Client:** Special Samples

**Client Sample ID:** WS-002

**Lab ID:** 2015-1649

**Collection Date:** 6/25/2015 8:50:00 AM

**Matrix:** Water

**Analyses**

**Dissolved Metals by EPA 200.8**

**EPA 200.8**

**Batch: 15080703 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Aluminum	<20	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	1.04	0.5	0.2		ug/L
Barium	10.9	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	13.6	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	5.14	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	<0.5	0.5	0.2		ug/L
Iron	<20	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	2.44	0.02	0.01		mg/L
Manganese	37.5	0.3	0.07		ug/L
Nickel	<0.5	0.5	0.15		ug/L
Potassium	2.24	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	2.81	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	3.04	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	1.05	1	0.3		ug/L
Hardness	22.9	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Aug 7 2015 4:04AM				

**Total Metals by EPA 200.8**

**EPA 200.8**

**Batch: 15081303 Run: 1**

**Result Reporting MDL Qual Unit**

		<b>Limit</b>		
Aluminum	<20	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	1.50	1	0.5	ug/L
Barium	18.2	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	5.33	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	<1	1	0.5	ug/L
Iron	291	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	2.40	0.1	0.1	mg/L
Manganese	317	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.26	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	3.58	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	<3	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Aug 11 2015 12:59AM			
Prep By				
Prep Date/Time				

**Client:** Special Samples

**Client Sample ID:** WS-005

**Lab ID:** 2015-1650

**Collection Date:** 6/25/2015 9:10:00 AM

**Matrix:** Water

**Analyses**

**Dissolved Metals by EPA 200.8**

**EPA 200.8**

**Batch: 15080703 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Aluminum	<20	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	1.41	0.5	0.2		ug/L
Barium	15.6	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	14.1	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	5.29	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	<0.5	0.5	0.2		ug/L
Iron	63.2	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	2.27	0.02	0.01		mg/L
Manganese	278	0.3	0.07		ug/L
Nickel	<0.5	0.5	0.15		ug/L
Potassium	2.35	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	3.44	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	2.92	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	4.64	1	0.3		ug/L
Hardness	22.5	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Aug 7 2015 4:11AM				

**Total Metals by EPA 200.8**

**EPA 200.8**

**Batch: 15081303 Run: 1**

**Result Reporting MDL Qual Unit**

		<b>Limit</b>		
Aluminum	27.2	20	20	ug/L
Antimony	<10	10	5	ug/L
Arsenic	2.14	1	0.5	ug/L
Barium	22.1	10	2.0	ug/L
Beryllium	<0.5	0.5	0.1	ug/L
Boron	<25	25	5.0	ug/L
Cadmium	<1	1	0.3	ug/L
Calcium	5.27	0.04	0.04	mg/L
Chromium	<1	1	0.3	ug/L
Cobalt	<1	1	0.5	ug/L
Copper	<1	1	0.5	ug/L
Iron	513	20	10.0	ug/L
Lead	<1	1	0.1	ug/L
Magnesium	2.36	0.1	0.1	mg/L
Manganese	479	1	0.2	ug/L
Nickel	<2.5	2.5	0.5	ug/L
Potassium	2.32	1	0.05	mg/L
Selenium	<2	2	0.5	ug/L
Silver	<5	5	1.0	ug/L
Sodium	3.73	0.04	0.02	mg/L
Thallium	<2.5	2.5	0.05	ug/L
Vanadium	<2.5	2.5	1.0	ug/L
Zinc	<3	3	2.0	ug/L
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Aug 11 2015 1:06AM			
Prep By				
Prep Date/Time				

<b>Client:</b> Special Samples	<b>Client Sample ID:</b> Metals Field Blank
<b>Lab ID:</b> 2015-1651	<b>Collection Date:</b> 6/25/2015 8:12:00 AM
<b>Matrix:</b> Water	

**Analyses**

<i>Dissolved Metals by EPA 200.8</i>	<i>EPA 200.8</i>	<i>Batch: 15080703 Run: 1</i>			
	<b><u>Result</u></b>	<b><u>Reporting Limit</u></b>	<b><u>MDL</u></b>	<b><u>Qual</u></b>	<b><u>Unit</u></b>
Aluminum	<20	20	20		ug/L
Antimony	<5	5	1.0		ug/L
Arsenic	<0.5	0.5	0.2		ug/L
Barium	<2	2	0.4		ug/L
Beryllium	<0.1	0.1	0.04		ug/L
Boron	<5	5	2.0		ug/L
Cadmium	<0.1	0.1	0.05		ug/L
Calcium	<0.03	0.03	0.03		mg/L
Chromium	<0.5	0.5	0.05		ug/L
Cobalt	<0.5	0.5	0.05		ug/L
Copper	<0.5	0.5	0.2		ug/L
Iron	<20	20	5.0		ug/L
Lead	<0.3	0.3	0.02		ug/L
Magnesium	<0.02	0.02	0.01		mg/L
Manganese	<0.3	0.3	0.07		ug/L
Nickel	<0.5	0.5	0.15		ug/L
Potassium	<0.02	0.02	0.01		mg/L
Selenium	<1	1	0.2		ug/L
Silicon Dioxide	<0.05	0.05	0.01		mg/L
Silver	<0.5	0.5	0.02		ug/L
Sodium	<0.02	0.02	0.01		mg/L
Thallium	<0.5	0.5	0.005		ug/L
Vanadium	<0.5	0.5	0.3		ug/L
Zinc	<1	1	0.3		ug/L
Hardness	<1	1	1.0		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	Aug 7 2015 4:17AM				



**Client:** Special Samples

**Client Sample ID:** WS-001

**Lab ID:** 2015-1648

**Collection Date:** 6/25/2015 8:09:00 AM

**Matrix:** Water

**Analyses**

**Oil and Grease**

**EPA1664**

**Batch: 15063003 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

**Semi-Volatiles by GC/MS**

**EPA 3510C/EPA 8270D**

**Batch: 15063001 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
2-Fluorophenol (% Recovery)	35.2	40-110			%
Nitrobenzene-d5 (% Recovery)	75.2	50-110			%
2-Fluorobiphenyl (% Recovery)	74.8	50-110			%
2,4,6-Tribromophenol (% Recovery)	78.6	40-110			%
Terphenyl-d14 (% Recovery)	82.8	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	<0.3	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	6/29/2015 1:34 PM			
Prep By	Ed Harris			
Prep Date/Time	06/29/2015 08:00			

**Client:** Special Samples

**Client Sample ID:** WS-002

**Lab ID:** 2015-1649

**Collection Date:** 6/25/2015 8:50:00 AM

**Matrix:** Water

**Analyses**

**Oil and Grease**

**EPA1664**

**Batch: 15063003 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

**Semi-Volatiles by GC/MS**

**EPA 3510C/EPA 8270D**

**Batch: 15063001 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
2-Fluorophenol (% Recovery)	37.7	40-110			%
Nitrobenzene-d5 (% Recovery)	68.9	50-110			%
2-Fluorobiphenyl (% Recovery)	69.1	50-110			%
2,4,6-Tribromophenol (% Recovery)	73.7	40-110			%
Terphenyl-d14 (% Recovery)	77.0	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	<0.3	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	6/29/2015 3:54 PM			
Prep By	Ed Harris			
Prep Date/Time	06/29/2015 08:00			

**Client:** Special Samples

**Client Sample ID:** WS-005

**Lab ID:** 2015-1650

**Collection Date:** 6/25/2015 9:10:00 AM

**Matrix:** Water

**Analyses**

**Oil and Grease**

**EPA1664**

**Batch: 15063003 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
Oil and Grease	<2.5	2.5	2.5		mg/L
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

**Semi-Volatiles by GC/MS**

**EPA 3510C/EPA 8270D**

**Batch: 15063001 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
2-Fluorophenol (% Recovery)	33.2	40-110			%
Nitrobenzene-d5 (% Recovery)	61.0	50-110			%
2-Fluorobiphenyl (% Recovery)	64.2	50-110			%
2,4,6-Tribromophenol (% Recovery)	77.5	40-110			%
Terphenyl-d14 (% Recovery)	82.7	50-110			%
Methyl Methanesulfonate	<0.2	0.2	100		ug/L
Ethyl methanesulfonate	<0.2	0.2	100		ug/L
Phenol	<0.2	0.2	100		ug/L
Aniline	<0.2	0.2	100		ug/L
Bis(2-chloroethyl)ether	<0.2	0.2	100		ug/L
2-Chlorophenol	<0.2	0.2	100		ug/L
1,3-Dichlorobenzene	<0.12	0.12	60		ug/L
1,4-Dichlorobenzene	<0.12	0.12	60		ug/L
Benzyl alcohol	<0.16	0.16	80		ug/L
1,2-Dichlorobenzene	<0.12	0.12	60		ug/L
2-Methylphenol	<0.1	0.1	50		ug/L
Acetophenone	<0.1	0.1	50		ug/L
4-Methylphenol	<0.1	0.1	50		ug/L
N-Nitrosodi-n-propylamine	<0.2	0.2	100		ug/L
Hexachloroethane	<0.2	0.2	100		ug/L
Nitrobenzene	<0.2	0.2	100		ug/L
N-Nitrosopiperidine	<0.2	0.2	100		ug/L
Isophorone	<0.1	0.1	50		ug/L

2-Nitrophenol	<0.3	0.3	150	ug/L
2,4-Dimethylphenol	<0.1	0.1	50	ug/L
Bis(2-chloroethoxy)methane	<0.2	0.2	100	ug/L
2,4-Dichlorophenol	<0.2	0.2	100	ug/L
1,2,4-Trichlorobenzene	<0.12	0.12	60	ug/L
Naphthalene	<0.08	0.08	40	ug/L
4-Chloroaniline	<0.1	0.1	50	ug/L
2,6-Dichlorophenol	<0.2	0.2	100	ug/L
Hexachlorobutadiene	<0.2	0.2	100	ug/L
N-Nitrosodibutylamine	<0.2	0.2	100	ug/L
4-Chloro-3-methylphenol	<0.16	0.16	80	ug/L
2-Methylnaphthalene	<0.1	0.1	50	ug/L
1,2,4,5-Tetrachlorobenzene	<0.1	0.1	50	ug/L
Hexachlorocyclopentadiene	<0.16	0.16	80	ug/L
2,4,6-Trichlorophenol	<0.2	0.2	100	ug/L
2,4,5-Trichlorophenol	<0.2	0.2	100	ug/L
2-Chloronaphthalene	<0.1	0.1	50	ug/L
1-Chloronaphthalene	<0.1	0.1	50	ug/L
2-Nitroaniline	<0.2	0.2	100	ug/L
Dimethyl phthalate	<0.2	0.2	100	ug/L
2,6-Dinitrotoluene	<0.2	0.2	100	ug/L
Acenaphthylene	<0.08	0.08	40	ug/L
3-Nitroaniline	<0.2	0.2	100	ug/L
Acenaphthene	<0.1	0.1	50	ug/L
2,4-Dinitrophenol	<4	4	2000	ug/L
Pentachlorobenzene	<0.12	0.12	60	ug/L
4-Nitrophenol	<2	2	1000	ug/L
Dibenzofuran	<0.1	0.1	50	ug/L
2,4-Dinitrotoluene	<0.2	0.2	100	ug/L
2,3,4,6-Tetrachlorophenol	<0.6	0.6	300	ug/L
Diethyl phthalate	<0.2	0.2	100	ug/L
Fluorene	<0.1	0.1	50	ug/L
4-Chlorophenyl phenyl ether	<0.1	0.1	50	ug/L
4-Nitroaniline	<0.2	0.2	100	ug/L
4,6-Dinitro-2-methylphenol	<6	6	3000	ug/L
Diphenylamine	<0.1	0.1	50	ug/L
Azobenzene	<0.08	0.08	40	ug/L
4-Bromophenyl phenyl ether	<0.2	0.2	100	ug/L
Hexachlorobenzene	<0.16	0.16	80	ug/L
Pentachlorophenol	<1	1	500	ug/L



Arkansas Department of Environmental Quality  
 5301 Northshore Drive  
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
 Ruehr@adeq.state.ar.us  
 501-682-0955

Pentachloronitrobenzene	<0.2	0.2	100	ug/L
Pronamide	<0.2	0.2	100	ug/L
Phenanthrene	<0.08	0.08	40	ug/L
Anthracene	<0.08	0.08	40	ug/L
Carbazole	<0.1	0.1	50	ug/L
Di-n-butyl phthalate	<0.2	0.2	100	ug/L
Fluoranthene	<0.08	0.08	40	ug/L
Pyrene	<0.08	0.08	40	ug/L
Dimethylaminoazobenzene	<0.2	0.2	100	ug/L
Butyl benzyl phthalate	<0.3	0.3	150	ug/L
Benzo (a) anthracene	<0.1	0.1	50	ug/L
Chrysene	<0.1	0.1	50	ug/L
Bis(2-ethylhexyl)phthalate	<0.3	0.3	150	ug/L
Di-n-octyl phthalate	<0.3	0.3	150	ug/L
Benzo (b) fluoranthene	<0.16	0.16	80	ug/L
7,12-Dimethylbenz (a) anthracene	<0.2	0.2	100	ug/L
Benzo (k) fluoranthene	<0.16	0.16	80	ug/L
Benzo (a) pyrene	<0.16	0.16	80	ug/L
3-Methylcholanthrene	<0.2	0.2	100	ug/L
Indeno (1,2,3-cd) pyrene	<0.2	0.2	100	ug/L
Dibenzo (a,h) anthracene	<0.16	0.16	80	ug/L
Benzo (g,h,i) perylene	<0.16	0.16	80	ug/L
Initial Volume	500			mL
Final Volume	1			mL
Dilution Factor	1			
Analyzed By	Ed Harris			
Analysis Date/Time	6/30/2015 9:45 AM			
Prep By	Ed Harris			
Prep Date/Time	06/29/2015 08:00			

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

**Client:** Special Samples

**Client Sample ID:** WS-001

**Lab ID:** 2015-1648

**Collection Date:** 6/25/2015 8:09:00 AM

**Matrix:** Water

**Analyses**

***Turbidity***

***EPA 180.1***

***Batch: 15062907 Run: 1***

	<b><u>Result</u></b>	<b><u>Reporting Limit</u></b>	<b><u>MDL</u></b>	<b><u>Qual</u></b>	<b><u>Unit</u></b>
Turbidity	8.61	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Kathryn Hattenhauer				
Analysis Date/Time	6/29/2015 11:22				

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

**Client:** Special Samples

**Client Sample ID:** WS-002

**Lab ID:** 2015-1649

**Collection Date:** 6/25/2015 8:50:00 AM

**Matrix:** Water

**Analyses**

***Turbidity***

***EPA 180.1***

***Batch: 15062907 Run: 1***

	<b><u>Result</u></b>	<b><u>Reporting Limit</u></b>	<b><u>MDL</u></b>	<b><u>Qual</u></b>	<b><u>Unit</u></b>
Turbidity	8.45	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Kathryn Hattenhauer				
Analysis Date/Time	6/29/2015 11:26				

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

**Client:** Special Samples

**Client Sample ID:** WS-005

**Lab ID:** 2015-1650

**Collection Date:** 6/25/2015 9:10:00 AM

**Matrix:** Water

**Analyses**

***Turbidity***

***EPA 180.1***

***Batch: 15062907 Run: 1***

	<b><u>Result</u></b>	<b><u>Reporting Limit</u></b>	<b><u>MDL</u></b>	<b><u>Qual</u></b>	<b><u>Unit</u></b>
Turbidity	11.6	0.02	0.02		NTU
Dilution Factor	1				
Analyzed By	Kathryn Hattenhauer				
Analysis Date/Time	6/29/2015 11:27				

**Client:** Special Samples

**Client Sample ID:** WS-001

**Lab ID:** 2015-1648

**Collection Date:** 6/25/2015 8:09:00 AM

**Matrix:** Water

**Analyses**

**Volatile Organics by GCMS**

**EPA 8260C**

**Batch: 15070101 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
<i>Dibromofluoromethane (% Recovery)</i>	107	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	102	70-130			%
<i>Toluene-d8 (% Recovery)</i>	105	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	117	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	6/30/2015 4:58 PM			

**Client:** Special Samples

**Client Sample ID:** WS-002

**Lab ID:** 2015-1649

**Collection Date:** 6/25/2015 8:50:00 AM

**Matrix:** Water

**Analyses**

**Volatile Organics by GCMS**

**EPA 8260C**

**Batch: 15070101 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
<i>Dibromofluoromethane (% Recovery)</i>	107	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	103	70-130			%
<i>Toluene-d8 (% Recovery)</i>	100	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	113	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	6/30/2015 4:08 PM			



**Client:** Special Samples

**Client Sample ID:** WS-005

**Lab ID:** 2015-1650

**Collection Date:** 6/25/2015 9:10:00 AM

**Matrix:** Water

**Analyses**

**Volatile Organics by GCMS**

**EPA 8260C**

**Batch: 15070101 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
<i>Dibromofluoromethane (% Recovery)</i>	107	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	102	70-130			%
<i>Toluene-d8 (% Recovery)</i>	101	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	121	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	6/30/2015 4:33 PM			

**Client:** Special Samples

**Client Sample ID:** Vol Trip Blank

**Lab ID:** 2015-1652

**Collection Date:** 6/25/2015 2:49:00 PM

**Matrix:** Water

**Analyses**

**Volatile Organics by GCMS**

**EPA 8260C**

**Batch: 15070101 Run: 1**

	<b>Result</b>	<b>Reporting Limit</b>	<b>MDL</b>	<b>Qual</b>	<b>Unit</b>
<i>Dibromofluoromethane (% Recovery)</i>	103	70-130			%
<i>1,2-Dichloroethane-d4 (% Recovery)</i>	98.4	70-130			%
<i>Toluene-d8 (% Recovery)</i>	99.9	70-130			%
<i>4-Bromofluorobenzene (% Recovery)</i>	114	70-130			%
Dichlorodifluoromethane	<1.12	1.12	1.12		ug/L
Chloromethane	<0.58	0.58	0.58		ug/L
Vinyl chloride	<0.82	0.82	0.82		ug/L
Bromomethane	<3.9	3.9	3.90		ug/L
Chloroethane	<2.68	2.68	2.68		ug/L
Trichlorofluoromethane	<0.51	0.51	0.51		ug/L
1,1-Dichloroethene	<0.43	0.43	0.43		ug/L
Acetone	<10.5	10.5	10.5		ug/L
Methylene chloride	<2.5	2.5	2.5		ug/L
Methyl tert-butyl ether	<0.83	0.83	0.83		ug/L
trans-1,2-Dichloroethene	<0.59	0.59	0.59		ug/L
1,1-Dichloroethane	<0.42	0.42	0.42		ug/L
Methyl ethyl ketone	<12.8	12.8	12.8		ug/L
cis-1,2-Dichloroethene	<1.15	1.15	1.15		ug/L
2,2-Dichloropropane	<0.81	0.81	0.81		ug/L
Bromochloromethane	<0.66	0.66	0.66		ug/L
Chloroform	<0.27	0.27	0.27		ug/L
1,1,1-Trichloroethane	<0.46	0.46	0.46		ug/L
1,1-Dichloropropene	<0.59	0.59	0.59		ug/L
Carbon tetrachloride	<0.6	0.6	0.6		ug/L
Benzene	<0.66	0.66	0.66		ug/L
1,2-Dichloroethane	<1.15	1.15	1.15		ug/L
Trichloroethene	<0.6	0.6	0.60		ug/L
1,2-Dichloropropane	<0.98	0.98	0.98		ug/L
Dibromomethane	<1.78	1.78	1.78		ug/L
Bromodichloromethane	<0.65	0.65	0.65		ug/L
cis-1,3-Dichloropropene	<0.86	0.86	0.86		ug/L
Methyl isobutyl ketone	<8.1	8.1	8.10		ug/L

Toluene	<0.57	0.57	0.57	ug/L
trans-1,3-Dichloropropene	<0.84	0.84	0.84	ug/L
1,1,2-Trichloroethane	<0.78	0.78	0.78	ug/L
2-Hexanone	<9.5	9.5	9.5	ug/L
Tetrachloroethene	<0.96	0.96	0.96	ug/L
1,3-Dichloropropane	<0.94	0.94	0.94	ug/L
Dibromochloromethane	<1.25	1.25	1.25	ug/L
1,2-Dibromoethane (EDB)	<0.68	0.68	0.68	ug/L
Chlorobenzene	<0.62	0.62	0.62	ug/L
Ethylbenzene	<0.51	0.51	0.51	ug/L
1,1,1,2-Tetrachloroethane	<0.57	0.57	0.57	ug/L
m,p-Xylene	<1.2	1.2	1.2	ug/L
o-Xylene	<0.5	0.5	0.5	ug/L
Styrene	<0.53	0.53	0.53	ug/L
Bromoform	<1.56	1.56	1.56	ug/L
Isopropylbenzene	<0.59	0.59	0.59	ug/L
1,1,1,2,2-Tetrachloroethane	<0.39	0.39	0.39	ug/L
1,2,3-Trichloropropane	<1.83	1.83	1.83	ug/L
n-Propylbenzene	<0.49	0.49	0.49	ug/L
Bromobenzene	<0.5	0.5	0.5	ug/L
1,3,5-Trimethylbenzene	<0.3	0.3	0.30	ug/L
2-Chlorotoluene	<0.66	0.66	0.66	ug/L
4-Chlorotoluene	<0.8	0.8	0.80	ug/L
tert-Butylbenzene	<0.85	0.85	0.85	ug/L
1,2,4-Trimethylbenzene	<0.46	0.46	0.46	ug/L
sec-Butylbenzene	<0.63	0.63	0.63	ug/L
p-Isopropyltoluene	<0.59	0.59	0.59	ug/L
1,3-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,4-Dichlorobenzene	<0.53	0.53	0.53	ug/L
n-Butylbenzene	<0.72	0.72	0.72	ug/L
1,2-Dichlorobenzene	<0.7	0.7	0.70	ug/L
1,2-Dibromo-3-chloropropane	<0.86	0.86	0.86	ug/L
1,2,4-Trichlorobenzene	<1.14	1.14	1.14	ug/L
Hexachlorobutadiene	<0.96	0.96	0.96	ug/L
Naphthalene	<1.53	1.53	1.53	ug/L
1,2,3-Trichlorobenzene	<1.3	1.3	1.3	ug/L
Dilution Factor	1			
Analyzed By	Jeff Ruehr			
Analysis Date/Time	6/30/2015 3:42 PM			

## Analytical Quality Control Results Report

<b>Batch: 15070101</b>	<b>VOA - water</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*Volatiles - water DUP*

*Run: 1*

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Dibromofluoromethane (% Recovery)	106 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	102 %			70 - 130	
Toluene-d8 (% Recovery)	101 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	116 %			70 - 130	
Dichlorodifluoromethane (RPD)	0 %				0 - 20
Chloromethane (RPD)	0 %				0 - 20
Vinyl chloride (RPD)	0 %				0 - 20
Bromomethane (RPD)	0 %				0 - 20
Chloroethane (RPD)	0 %				0 - 20
Trichlorofluoromethane (RPD)	0 %				0 - 20
1,1-Dichloroethene (RPD)	0 %				0 - 20
Acetone (RPD)	25.0 %				0 - 20
Methylene chloride (RPD)	0 %				0 - 20
Methyl tert-butyl ether (RPD)	0 %				0 - 20
trans-1,2-Dichloroethene (RPD)	0 %				0 - 20
1,1-Dichloroethane (RPD)	0 %				0 - 20
Methyl ethyl ketone (RPD)	0 %				0 - 20
cis-1,2-Dichloroethene (RPD)	0 %				0 - 20
2,2-Dichloropropane (RPD)	0 %				0 - 20
Bromochloromethane (RPD)	0 %				0 - 20
Chloroform (RPD)	0 %				0 - 20
1,1,1-Trichloroethane (RPD)	0 %				0 - 20
1,1-Dichloropropene (RPD)	0 %				0 - 20
Carbon tetrachloride (RPD)	0 %				0 - 20
Benzene (RPD)	0 %				0 - 20
1,2-Dichloroethane (RPD)	0 %				0 - 20
Trichloroethene (RPD)	0 %				0 - 20
1,2-Dichloropropane (RPD)	0 %				0 - 20
Dibromomethane (RPD)	0 %				0 - 20
Bromodichloromethane (RPD)	0 %				0 - 20
cis-1,3-Dichloropropene (RPD)	0 %				0 - 20

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

Methyl isobutyl ketone (RPD)	0 %	0 - 20
Toluene (RPD)	0 %	0 - 20
trans-1,3-Dichloropropene (RPD)	0 %	0 - 20
1,1,2-Trichloroethane (RPD)	0 %	0 - 20
2-Hexanone (RPD)	0 %	0 - 20
Tetrachloroethene (RPD)	0 %	0 - 20
1,3-Dichloropropane (RPD)	0 %	0 - 20
Dibromochloromethane (RPD)	0 %	0 - 20
1,2-Dibromoethane (EDB) (RPD)	0 %	0 - 20
Chlorobenzene (RPD)	0 %	0 - 20
Ethylbenzene (RPD)	0 %	0 - 20
1,1,1,2-Tetrachloroethane (RPD)	0 %	0 - 20
m,p-Xylene (RPD)	0 %	0 - 20
o-Xylene (RPD)	0 %	0 - 20
Styrene (RPD)	0 %	0 - 20
Bromoform (RPD)	0 %	0 - 20
Isopropylbenzene (RPD)	0 %	0 - 20
1,1,2,2-Tetrachloroethane (RPD)	0 %	0 - 20
1,2,3-Trichloropropane (RPD)	0 %	0 - 20
n-Propylbenzene (RPD)	0 %	0 - 20
Bromobenzene (RPD)	0 %	0 - 20
1,3,5-Trimethylbenzene (RPD)	0 %	0 - 20
2-Chlorotoluene (RPD)	0 %	0 - 20
4-Chlorotoluene (RPD)	0 %	0 - 20
tert-Butylbenzene (RPD)	0 %	0 - 20
1,2,4-Trimethylbenzene (RPD)	0 %	0 - 20
sec-Butylbenzene (RPD)	0 %	0 - 20
p-Isopropyltoluene (RPD)	0 %	0 - 20
1,3-Dichlorobenzene (RPD)	0 %	0 - 20
1,4-Dichlorobenzene (RPD)	0 %	0 - 20
n-Butylbenzene (RPD)	0 %	0 - 20
1,2-Dichlorobenzene (RPD)	0 %	0 - 20
1,2-Dibromo-3-chloropropane (RPD)	0 %	0 - 20
1,2,4-Trichlorobenzene (RPD)	0 %	0 - 20
Hexachlorobutadiene (RPD)	0 %	0 - 20
Naphthalene (RPD)	0 %	0 - 20
1,2,3-Trichlorobenzene (RPD)	0 %	0 - 20
Dilution Factor	1	
Analyzed By	Jeff Ruehr	
Analysis Date/Time	6/30/2015 5:23 PM	

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

**Volatiles - water MS**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Dibromofluoromethane (% Recovery)	108 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	106 %			70 - 130	
Toluene-d8 (% Recovery)	104 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	119 %			70 - 130	
1,1-Dichloroethene (% Recovery)	114 %			70 - 130	
Benzene (% Recovery)	100 %			70 - 130	
Trichloroethene (% Recovery)	91.0 %			70 - 130	
Toluene (% Recovery)	91.5 %			70 - 130	
Chlorobenzene (% Recovery)	96.1 %			70 - 130	
Dilution Factor	1				
Analyzed By	Jeff Ruehr				
Analysis Date/Time	6/30/2015 5:49 PM				

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

**Volatiles - water MSD**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Dibromofluoromethane (% Recovery)	111 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	107 %			70 - 130	
Toluene-d8 (% Recovery)	102 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	117 %			70 - 130	
1,1-Dichloroethene (% Recovery)	116 %			70 - 130	
1,1-Dichloroethene (RPD)	2.2 %				0 - 20
Benzene (RPD)	0.1 %				0 - 20
Benzene (% Recovery)	100 %			70 - 130	
Trichloroethene (% Recovery)	90.0 %			70 - 130	
Trichloroethene (RPD)	1.0 %				0 - 20
Toluene (RPD)	0.6 %				0 - 20
Toluene (% Recovery)	92.1 %			70 - 130	
Chlorobenzene (RPD)	0.9 %				0 - 20
Chlorobenzene (% Recovery)	96.9 %			70 - 130	
Dilution Factor	1				
Analyzed By	Jeff Ruehr				
Analysis Date/Time	6/30/2015 6:14 PM				

<b>LCS</b>	<b>LIMS ID: 15070101-LCS-01</b>
------------	---------------------------------

**Volatiles - water LCS**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Dibromofluoromethane (% Recovery)	99.6 %			70 - 130	
1,2-Dichloroethane-d4 (% Recovery)	100 %			70 - 130	
Toluene-d8 (% Recovery)	103 %			70 - 130	
4-Bromofluorobenzene (% Recovery)	111 %			70 - 130	
Dichlorodifluoromethane (% Recovery)	122 %			60 - 130	
Chloromethane (% Recovery)	95.1 %			60 - 130	
Vinyl chloride (% Recovery)	102 %			60 - 130	
Bromomethane (% Recovery)	81.3 %			60 - 130	
Chloroethane (% Recovery)	79.6 %			60 - 130	
Trichlorofluoromethane (% Recovery)	103 %			60 - 130	
1,1-Dichloroethene (% Recovery)	103 %			60 - 130	
Acetone (% Recovery)	99.0 %			60 - 130	
Methylene chloride (% Recovery)	85.4 %			60 - 130	
Methyl tert-butyl ether (% Recovery)	99.4 %			60 - 130	
trans-1,2-Dichloroethene (% Recovery)	103 %			60 - 130	
1,1-Dichloroethane (% Recovery)	100 %			60 - 130	
Methyl ethyl ketone (% Recovery)	102 %			60 - 130	
cis-1,2-Dichloroethene (% Recovery)	103 %			60 - 130	
2,2-Dichloropropane (% Recovery)	106 %			60 - 130	
Bromochloromethane (% Recovery)	101 %			60 - 130	
Chloroform (% Recovery)	99.0 %			60 - 130	
1,1,1-Trichloroethane (% Recovery)	100 %			60 - 130	
1,1-Dichloropropene (% Recovery)	104 %			60 - 130	
Carbon tetrachloride (% Recovery)	103 %			60 - 130	
Benzene (% Recovery)	99.3 %			60 - 130	
1,2-Dichloroethane (% Recovery)	99.8 %			60 - 130	
Trichloroethene (% Recovery)	101 %			60 - 130	
1,2-Dichloropropane (% Recovery)	103 %			60 - 130	
Dibromomethane (% Recovery)	98.0 %			60 - 130	
Bromodichloromethane (% Recovery)	99.3 %			60 - 130	
cis-1,3-Dichloropropene (% Recovery)	100 %			60 - 130	
Methyl isobutyl ketone (% Recovery)	104 %			60 - 130	
Toluene (% Recovery)	104 %			60 - 130	
trans-1,3-Dichloropropene (% Recovery)	99.8 %			60 - 130	
1,1,2-Trichloroethane (% Recovery)	99.0 %			60 - 130	
2-Hexanone (% Recovery)	98.3 %			60 - 130	



Tetrachloroethene (% Recovery)	107 %	60 - 130
1,3-Dichloropropane (% Recovery)	100 %	60 - 130
Dibromochloromethane (% Recovery)	99.5 %	60 - 130
1,2-Dibromoethane (EDB) (% Recovery)	98.5 %	60 - 130
Chlorobenzene (% Recovery)	105 %	60 - 130
Ethylbenzene (% Recovery)	103 %	60 - 130
1,1,1,2-Tetrachloroethane (% Recovery)	102 %	60 - 130
m,p-Xylene (% Recovery)	106 %	60 - 130
o-Xylene (% Recovery)	105 %	60 - 130
Styrene (% Recovery)	99.4 %	60 - 130
Bromoform (% Recovery)	98.8 %	60 - 130
Isopropylbenzene (% Recovery)	110 %	60 - 130
1,1,2,2-Tetrachloroethane (% Recovery)	99.7 %	60 - 130
1,2,3-Trichloropropane (% Recovery)	103 %	60 - 130
n-Propylbenzene (% Recovery)	107 %	60 - 130
Bromobenzene (% Recovery)	104 %	60 - 130
1,3,5-Trimethylbenzene (% Recovery)	114 %	60 - 130
2-Chlorotoluene (% Recovery)	111 %	60 - 130
4-Chlorotoluene (% Recovery)	110 %	60 - 130
tert-Butylbenzene (% Recovery)	103 %	60 - 130
1,2,4-Trimethylbenzene (% Recovery)	110 %	60 - 130
sec-Butylbenzene (% Recovery)	104 %	60 - 130
p-Isopropyltoluene (% Recovery)	112 %	60 - 130
1,3-Dichlorobenzene (% Recovery)	109 %	60 - 130
1,4-Dichlorobenzene (% Recovery)	108 %	60 - 130
n-Butylbenzene (% Recovery)	111 %	60 - 130
1,2-Dichlorobenzene (% Recovery)	106 %	60 - 130
1,2-Dibromo-3-chloropropane (% Recovery)	99.9 %	60 - 130
1,2,4-Trichlorobenzene (% Recovery)	112 %	60 - 130
Hexachlorobutadiene (% Recovery)	116 %	60 - 130
Naphthalene (% Recovery)	106 %	60 - 130
1,2,3-Trichlorobenzene (% Recovery)	112 %	60 - 130
Dilution Factor	1	
Analyzed By	Jeff Ruehr	
Analysis Date/Time	6/30/2015 14:01	

## Analytical Quality Control Results Report

<b>Batch: 15063001</b>	<b>Semi-VOA water (Prep)</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*Semi Volatiles - water DUP*

*Run: 1*

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	06/29/2015 08:00				
2-Fluorophenol (% Recovery)	37.3 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	82.3 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	82.7 %			40 - 110	
2,4,6-Tribromophenol (% Recovery)	89.1 %			40 - 110	
Terphenyl-d14 (% Recovery)	85.4 %			40 - 125	
Methyl Methanesulfonate (RPD)	0 %				0 - 40
Ethyl methanesulfonate (RPD)	0 %				0 - 40
Phenol (RPD)	1.58 %				0 - 40
Aniline (RPD)	15.1 %				0 - 40
Bis(2-chloroethyl)ether (RPD)	200 %				0 - 40
2-Chlorophenol (RPD)	0 %				0 - 40
1,3-Dichlorobenzene (RPD)	0 %				0 - 40
1,4-Dichlorobenzene (RPD)	0 %				0 - 40
Benzyl alcohol (RPD)	0.19 %				0 - 40
1,2-Dichlorobenzene (RPD)	0 %				0 - 40
2-Methylphenol (RPD)	0 %				0 - 40
Acetophenone (RPD)	200 %				0 - 40
4-Methylphenol (RPD)	0 %				0 - 40
N-Nitrosodi-n-propylamine (RPD)	0 %				0 - 40
Hexachloroethane (RPD)	0 %				0 - 40
Nitrobenzene (RPD)	0 %				0 - 40
N-Nitrosopiperidine (RPD)	0 %				0 - 40
Isophorone (RPD)	0 %				0 - 40
2-Nitrophenol (RPD)	0 %				0 - 40
2,4-Dimethylphenol (RPD)	0 %				0 - 40
Bis(2-chloroethoxy)methane (RPD)	0 %				0 - 40
2,4-Dichlorophenol (RPD)	0 %				0 - 40
1,2,4-Trichlorobenzene (RPD)	0 %				0 - 40
Naphthalene (RPD)	0 %				0 - 40

4-Chloroaniline (RPD)	0 %	0 - 40
2,6-Dichlorophenol (RPD)	0 %	0 - 40
Hexachlorobutadiene (RPD)	0 %	0 - 40
N-Nitrosodibutylamine (RPD)	0 %	0 - 40
4-Chloro-3-methylphenol (RPD)	0 %	0 - 40
2-Methylnaphthalene (RPD)	0 %	0 - 40
1,2,4,5-Tetrachlorobenzene (RPD)	0 %	0 - 40
Hexachlorocyclopentadiene (RPD)	0 %	0 - 40
2,4,6-Trichlorophenol (RPD)	0 %	0 - 40
2,4,5-Trichlorophenol (RPD)	0 %	0 - 40
2-Chloronaphthalene (RPD)	0 %	0 - 40
1-Chloronaphthalene (RPD)	0 %	0 - 40
2-Nitroaniline (RPD)	0 %	0 - 40
Dimethyl phthalate (RPD)	0 %	0 - 40
2,6-Dinitrotoluene (RPD)	0 %	0 - 40
Acenaphthylene (RPD)	0 %	0 - 40
3-Nitroaniline (RPD)	200 %	0 - 40
Acenaphthene (RPD)	0 %	0 - 40
2,4-Dinitrophenol (RPD)	0 %	0 - 40
Pentachlorobenzene (RPD)	0 %	0 - 40
4-Nitrophenol (RPD)	200 %	0 - 40
Dibenzofuran (RPD)	0 %	0 - 40
2,4-Dinitrotoluene (RPD)	0 %	0 - 40
2,3,4,6-Tetrachlorophenol (RPD)	0 %	0 - 40
Diethyl phthalate (RPD)	0 %	0 - 40
Fluorene (RPD)	0 %	0 - 40
4-Chlorophenyl phenyl ether (RPD)	0 %	0 - 40
4-Nitroaniline (RPD)	0 %	0 - 40
4,6-Dinitro-2-methylphenol (RPD)	0 %	0 - 40
Diphenylamine (RPD)	0 %	0 - 40
Azobenzene (RPD)	0 %	0 - 40
4-Bromophenyl phenyl ether (RPD)	0 %	0 - 40
Hexachlorobenzene (RPD)	0 %	0 - 40
Pentachlorophenol (RPD)	200 %	0 - 40
Pentachloronitrobenzene (RPD)	0 %	0 - 40
Pronamide (RPD)	0 %	0 - 40
Phenanthrene (RPD)	0 %	0 - 40
Anthracene (RPD)	0 %	0 - 40
Carbazole (RPD)	0 %	0 - 40
Di-n-butyl phthalate (RPD)	1.37 %	0 - 40

Fluoranthene (RPD)	0 %	0 - 40
Pyrene (RPD)	0 %	0 - 40
Dimethylaminoazobenzene (RPD)	0 %	0 - 40
Butyl benzyl phthalate (RPD)	0 %	0 - 40
Benzo (a) anthracene (RPD)	0 %	0 - 40
Chrysene (RPD)	0 %	0 - 40
Bis(2-ethylhexyl)phthalate (RPD)	47.6 %	0 - 40
Di-n-octyl phthalate (RPD)	17.6 %	0 - 40
Benzo (b) fluoranthene (RPD)	0 %	0 - 40
7,12-Dimethylbenz (a) anthracene (RPD)	0 %	0 - 40
Benzo (k) fluoranthene (RPD)	0 %	0 - 40
Benzo (a) pyrene (RPD)	200 %	0 - 40
3-Methylcholanthrene (RPD)	0 %	0 - 40
Indeno (1,2,3-cd) pyrene (RPD)	0 %	0 - 40
Dibenzo (a,h) anthracene (RPD)	0 %	0 - 40
Benzo (g,h,i) perylene (RPD)	0 %	0 - 40
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	6/29/2015 2:30 PM	

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

**Semi Volatiles - water MS**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	06/29/2015 08:00				
2-Fluorophenol (% Recovery)	39.6 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	73.1 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	79.1 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	89.0 %			40 - 125	
Terphenyl-d14 (% Recovery)	87.6 %			40 - 125	
Phenol (% Recovery)	28.1 %			25 - 125	
2-Chlorophenol (% Recovery)	62.7 %			25 - 125	
1,4-Dichlorobenzene (% Recovery)	69.9 %			25 - 125	
N-Nitrosodi-n-propylamine (% Recovery)	72.9 %			25 - 125	
1,2,4-Trichlorobenzene (% Recovery)	67.3 %			25 - 125	
4-Chloro-3-methylphenol (% Recovery)	65.4 %			25 - 125	
Acenaphthene (% Recovery)	67.9 %			25 - 125	

4-Nitrophenol (% Recovery)	33.5 %	25 - 125
2,4-Dinitrotoluene (% Recovery)	75.2 %	25 - 125
Pentachlorophenol (% Recovery)	107 %	25 - 125
Pyrene (% Recovery)	90.1 %	25 - 125
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	6/29/2015 2:58 PM	

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

**Semi Volatiles - water MSD**

**Run: 1**

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	06/29/2015 08:00				
2-Fluorophenol (% Recovery)	43.5 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	76.1 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	75.7 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	80.9 %			40 - 125	
Terphenyl-d14 (% Recovery)	85.1 %			40 - 125	
Phenol (% Recovery)	32.1 %			25 - 125	
Phenol (RPD)	13.2 %				0 - 40
2-Chlorophenol (% Recovery)	68.5 %			25 - 125	
2-Chlorophenol (RPD)	8.7 %				0 - 40
1,4-Dichlorobenzene (RPD)	3.1 %				0 - 40
1,4-Dichlorobenzene (% Recovery)	72.1 %			25 - 125	
N-Nitrosodi-n-propylamine (% Recovery)	74.8 %			25 - 125	
N-Nitrosodi-n-propylamine (RPD)	2.5 %				0 - 40
1,2,4-Trichlorobenzene (% Recovery)	68.7 %			25 - 125	
1,2,4-Trichlorobenzene (RPD)	2.0 %				0 - 40
4-Chloro-3-methylphenol (% Recovery)	71.4 %			25 - 125	
4-Chloro-3-methylphenol (RPD)	8.8 %				0 - 40
Acenaphthene (% Recovery)	69.4 %			25 - 125	
Acenaphthene (RPD)	2.2 %				0 - 40
4-Nitrophenol (% Recovery)	35.7 %			25 - 125	
4-Nitrophenol (RPD)	6.4 %				0 - 40
2,4-Dinitrotoluene (% Recovery)	77.6 %			25 - 125	
2,4-Dinitrotoluene (RPD)	3.1 %				0 - 40
Pentachlorophenol (% Recovery)	101 %			25 - 125	

Pentachlorophenol (RPD)	5.8 %	0 - 40
Pyrene (% Recovery)	85.3 %	25 - 125
Pyrene (RPD)	5.5 %	0 - 40
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	6/29/2015 3:26 PM	

<b>MB</b>	<b>LIMS ID: 15063001-MB-01</b>
-----------	--------------------------------

**Semi Volatiles - water MB**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	06/29/2015 08:00				
2-Fluorophenol (% Recovery)	40.1 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	85.3 %			40 - 125	
2-Fluorobiphenyl (% Recovery)	87.2 %			40 - 125	
2,4,6-Tribromophenol (% Recovery)	83.0 %			40 - 125	
Terphenyl-d14 (% Recovery)	122 %			40 - 125	
Methyl Methanesulfonate	<0.2 ug/L	0.2	0.2		
Ethyl methanesulfonate	<0.2 ug/L	0.2	0.2		
Phenol	<0.2 ug/L	0.2	0.2		
Aniline	<0.2 ug/L	0.2	0.2		
Bis(2-chloroethyl)ether	<0.2 ug/L	0.2	0.2		
2-Chlorophenol	<0.2 ug/L	0.2	0.2		
1,3-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
1,4-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
Benzyl alcohol	<0.16 ug/L	0.16	0.16		
1,2-Dichlorobenzene	<0.12 ug/L	0.12	0.12		
2-Methylphenol	<0.1 ug/L	0.1	0.1		
Acetophenone	<0.1 ug/L	0.1	0.1		
4-Methylphenol	<0.1 ug/L	0.1	0.1		
N-Nitrosodi-n-propylamine	<0.2 ug/L	0.2	0.2		
Hexachloroethane	<0.2 ug/L	0.2	0.2		
Nitrobenzene	<0.2 ug/L	0.2	0.2		
N-Nitrosopiperidine	<0.2 ug/L	0.2	0.2		
Isophorone	<0.1 ug/L	0.1	0.1		
2-Nitrophenol	<0.3 ug/L	0.3	0.3		
2,4-Dimethylphenol	<0.1 ug/L	0.1	0.1		

Bis(2-chloroethoxy)methane	<0.2 ug/L	0.2	0.2
2,4-Dichlorophenol	<0.2 ug/L	0.2	0.2
1,2,4-Trichlorobenzene	<0.12 ug/L	0.12	0.12
Naphthalene	<0.08 ug/L	0.08	0.08
4-Chloroaniline	<0.1 ug/L	0.1	0.1
2,6-Dichlorophenol	<0.2 ug/L	0.2	0.2
Hexachlorobutadiene	<0.2 ug/L	0.2	0.2
N-Nitrosodibutylamine	<0.2 ug/L	0.2	0.2
4-Chloro-3-methylphenol	<0.16 ug/L	0.16	0.16
2-Methylnaphthalene	<0.12 ug/L	0.12	0.12
1,2,4,5-Tetrachlorobenzene	<0.1 ug/L	0.1	0.1
Hexachlorocyclopentadiene	<0.16 ug/L	0.16	0.16
2,4,6-Trichlorophenol	<0.2 ug/L	0.2	0.2
2,4,5-Trichlorophenol	<0.2 ug/L	0.2	0.2
2-Chloronaphthalene	<0.1 ug/L	0.1	0.1
1-Chloronaphthalene	<0.1 ug/L	0.1	0.1
2-Nitroaniline	<0.2 ug/L	0.2	0.2
Dimethyl phthalate	<0.2 ug/L	0.2	0.2
2,6-Dinitrotoluene	<0.2 ug/L	0.2	0.2
Acenaphthylene	<0.08 ug/L	0.08	0.08
3-Nitroaniline	<0.2 ug/L	0.2	0.2
Acenaphthene	<0.1 ug/L	0.1	0.1
2,4-Dinitrophenol	<4 ug/L	4	4
Pentachlorobenzene	<0.12 ug/L	0.12	0.12
4-Nitrophenol	<2 ug/L	2	2
Dibenzofuran	<0.1 ug/L	0.1	0.1
2,4-Dinitrotoluene	<0.2 ug/L	0.2	0.2
2,3,4,6-Tetrachlorophenol	<0.6 ug/L	0.6	0.6
Diethyl phthalate	<0.2 ug/L	0.2	0.2
Fluorene	<0.1 ug/L	0.1	0.1
4-Chlorophenyl phenyl ether	<0.1 ug/L	0.1	0.1
4-Nitroaniline	<0.2 ug/L	0.2	0.2
4,6-Dinitro-2-methylphenol	<6 ug/L	6	6
Diphenylamine	<0.1 ug/L	0.1	0.1
Azobenzene	<0.08 ug/L	0.08	0.08
4-Bromophenyl phenyl ether	<0.2 ug/L	0.2	0.2
Hexachlorobenzene	<0.16 ug/L	0.16	0.16
Pentachlorophenol	<1 ug/L	1	1
Pentachloronitrobenzene	<0.2 ug/L	0.2	0.2
Pronamide	<0.2 ug/L	0.2	0.2

Phenanthrene	<0.08 ug/L	0.08	0.08
Anthracene	<0.08 ug/L	0.08	0.08
Carbazole	<0.1 ug/L	0.1	0.1
Di-n-butyl phthalate	0.459 ug/L	0.2	0.2
Fluoranthene	<0.08 ug/L	0.08	0.08
Pyrene	<0.08 ug/L	0.08	0.08
Dimethylaminoazobenzene	<0.2 ug/L	0.2	0.2
Butyl benzyl phthalate	<0.3 ug/L	0.3	0.3
Benzo (a) anthracene	<0.2 ug/L	0.2	0.2
Chrysene	<0.1 ug/L	0.1	0.1
Bis(2-ethylhexyl)phthalate	<0.3 ug/L	0.3	0.3
Di-n-octyl phthalate	<0.3 ug/L	0.3	0.3
Benzo (b) fluoranthene	<0.16 ug/L	0.16	0.16
7,12-Dimethylbenz (a) anthracene	<0.2 ug/L	0.2	0.2
Benzo (k) fluoranthene	<0.16 ug/L	0.16	0.16
Benzo (a) pyrene	<0.16 ug/L	0.16	0.16
3-Methylcholanthrene	<0.2 ug/L	0.2	0.2
Indeno (1,2,3-cd) pyrene	<0.2 ug/L	0.2	0.2
Dibenzo (a,h) anthracene	<0.16 ug/L	0.16	0.16
Benzo (g,h,i) perylene	<0.16 ug/L	0.16	0.16
Dilution Factor	1		
Analyzed By	Ed Harris		
Analysis Date/Time	6/29/2015 12:38		

**LCS** **LIMS ID: 15063001-LCS-01**

**Semi Volatiles - water LCS**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Initial Volume	500 mL				
Final Volume	1 mL				
Prep By	Ed Harris				
Prep Date/Time	06/29/2015 08:00				
2-Fluorophenol (% Recovery)	33.4 %			40 - 110	
Nitrobenzene-d5 (% Recovery)	77.2 %			50 - 125	
2-Fluorobiphenyl (% Recovery)	78.6 %			50 - 125	
2,4,6-Tribromophenol (% Recovery)	84.6 %			40 - 125	
Terphenyl-d14 (% Recovery)	88.4 %			50 - 125	
Methyl Methanesulfonate (% Recovery)	60.5 %			50 - 150	
Ethyl methanesulfonate (% Recovery)	71.8 %			50 - 150	
Phenol (% Recovery)	30.0 %			50 - 150	



Aniline (% Recovery)	42.5 %	50 - 150
Bis(2-chloroethyl)ether (% Recovery)	78.6 %	50 - 150
2-Chlorophenol (% Recovery)	70.7 %	50 - 150
1,3-Dichlorobenzene (% Recovery)	58.8 %	50 - 150
1,4-Dichlorobenzene (% Recovery)	62.4 %	50 - 150
Benzyl alcohol (% Recovery)	57.4 %	50 - 150
1,2-Dichlorobenzene (% Recovery)	64.9 %	50 - 150
2-Methylphenol (% Recovery)	62.9 %	50 - 150
Acetophenone (% Recovery)	79.7 %	50 - 150
4-Methylphenol (% Recovery)	53.3 %	50 - 150
N-Nitrosodi-n-propylamine (% Recovery)	87.6 %	50 - 150
Hexachloroethane (% Recovery)	59.2 %	50 - 150
Nitrobenzene (% Recovery)	83.4 %	50 - 150
N-Nitrosopiperidine (% Recovery)	89.4 %	50 - 150
Isophorone (% Recovery)	88.6 %	50 - 150
2-Nitrophenol (% Recovery)	79.4 %	50 - 150
2,4-Dimethylphenol (% Recovery)	7.20 %	50 - 150
Bis(2-chloroethoxy)methane (% Recovery)	77.7 %	50 - 150
2,4-Dichlorophenol (% Recovery)	72.6 %	50 - 150
1,2,4-Trichlorobenzene (% Recovery)	71.1 %	50 - 150
Naphthalene (% Recovery)	77.9 %	50 - 150
4-Chloroaniline (% Recovery)	59.1 %	50 - 150
2,6-Dichlorophenol (% Recovery)	78.8 %	50 - 150
Hexachlorobutadiene (% Recovery)	59.0 %	50 - 150
N-Nitrosodibutylamine (% Recovery)	89.8 %	50 - 150
4-Chloro-3-methylphenol (% Recovery)	78.3 %	50 - 150
2-Methylnaphthalene (% Recovery)	78.6 %	50 - 150
1,2,4,5-Tetrachlorobenzene (% Recovery)	72.0 %	50 - 150
Hexachlorocyclopentadiene (% Recovery)	84.1 %	50 - 150
2,4,6-Trichlorophenol (% Recovery)	83.0 %	50 - 150
2,4,5-Trichlorophenol (% Recovery)	74.1 %	50 - 150
2-Chloronaphthalene (% Recovery)	78.5 %	50 - 150
1-Chloronaphthalene (% Recovery)	82.3 %	50 - 150
2-Nitroaniline (% Recovery)	79.5 %	50 - 150
Dimethyl phthalate (% Recovery)	79.8 %	50 - 150
2,6-Dinitrotoluene (% Recovery)	86.8 %	50 - 150
Acenaphthylene (% Recovery)	81.5 %	50 - 150
3-Nitroaniline (% Recovery)	79.5 %	50 - 150
Acenaphthene (% Recovery)	83.2 %	50 - 150
2,4-Dinitrophenol (% Recovery)	79.2 %	50 - 150

Pentachlorobenzene (% Recovery)	71.7 %	50 - 150
4-Nitrophenol (% Recovery)	29.0 %	50 - 150
Dibenzofuran (% Recovery)	82.5 %	50 - 150
2,4-Dinitrotoluene (% Recovery)	83.4 %	50 - 150
2,3,4,6-Tetrachlorophenol (% Recovery)	81.8 %	50 - 150
Diethyl phthalate (% Recovery)	91.7 %	50 - 150
Fluorene (% Recovery)	81.2 %	50 - 150
4-Chlorophenyl phenyl ether (% Recovery)	80.4 %	50 - 150
4-Nitroaniline (% Recovery)	70.6 %	50 - 150
4,6-Dinitro-2-methylphenol (% Recovery)	91.2 %	50 - 150
Diphenylamine (% Recovery)	78.8 %	50 - 150
Azobenzene (% Recovery)	79.4 %	50 - 150
4-Bromophenyl phenyl ether (% Recovery)	84.6 %	50 - 150
Hexachlorobenzene (% Recovery)	79.8 %	50 - 150
Pentachlorophenol (% Recovery)	80.9 %	50 - 150
Pentachloronitrobenzene (% Recovery)	85.6 %	50 - 150
Pronamide (% Recovery)	91.1 %	50 - 150
Phenanthrene (% Recovery)	83.2 %	50 - 150
Anthracene (% Recovery)	83.0 %	50 - 150
Carbazole (% Recovery)	99.6 %	50 - 150
Di-n-butyl phthalate (% Recovery)	118 %	50 - 150
Fluoranthene (% Recovery)	93.2 %	50 - 150
Pyrene (% Recovery)	87.8 %	50 - 150
Dimethylaminoazobenzene (% Recovery)	87.7 %	50 - 150
Butyl benzyl phthalate (% Recovery)	94.7 %	50 - 150
Benzo (a) anthracene (% Recovery)	88.7 %	50 - 150
Chrysene (% Recovery)	90.3 %	50 - 150
Bis(2-ethylhexyl)phthalate (% Recovery)	97.0 %	50 - 150
Di-n-octyl phthalate (% Recovery)	98.3 %	50 - 150
Benzo (b) fluoranthene (% Recovery)	94.0 %	50 - 150
7,12-Dimethylbenz (a) anthracene (% Recovery)	78.3 %	50 - 150
Benzo (k) fluoranthene (% Recovery)	95.0 %	50 - 150
Benzo (a) pyrene (% Recovery)	91.4 %	50 - 150
3-Methylcholanthrene (% Recovery)	81.9 %	50 - 150
Indeno (1,2,3-cd) pyrene (% Recovery)	92.3 %	50 - 150
Dibenzo (a,h) anthracene (% Recovery)	92.3 %	50 - 150
Benzo (g,h,i) perylene (% Recovery)	102 %	50 - 150
Dilution Factor	1	
Analyzed By	Ed Harris	
Analysis Date/Time	6/29/2015 13:06	

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

## Analytical Quality Control Results Report

<b>Batch: 15063003</b>	<b>Oil and Grease - water</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*Oil and Grease - water DUP*

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (RPD)	200 %				0 - 20
Oil and Grease	<2.5 mg/L	2.5	2.5		
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

*Oil and Grease - water MS*

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (% Recovery)	106 %			70 - 130	
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>
---------------	---------------------------

*Oil and Grease - water MSD*

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease (% Recovery)	96.5 %			70 - 130	
Oil and Grease (RPD)	9.8 %				0 - 20
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

<b>MB</b>	<b>LIMS ID: 15063003-MB-01</b>
-----------	--------------------------------

*Oil and Grease - water MB*

Run: 1

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Oil and Grease	<2.5 mg/L	2.5	2.5		
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

Arkansas Department of Environmental Quality  
5301 Northshore Drive  
North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
Ruehr@adeq.state.ar.us  
501-682-0955

**LCS** **LIMS ID: 15063003-LCS-01**

**Oil and Grease - water LCS**

**Run: 1**

<b>Parameter</b>	<b>Result</b>	<b>DL</b>	<b>RL</b>	<b>Accuracy Control</b>	<b>Precision Control</b>
Oil and Grease (% Recovery)	89.8 %			70 - 130	
Dilution Factor	1				
Analyzed By	Robert Graddy				
Analysis Date/Time	6-30-2015 15:15				

## Analytical Quality Control Results Report

<b>Batch: 15081303</b>	<b>ICP Metals - water (total)</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*ICP Metals - water (Total) DUP*

*Run: 1*

Parameter	Result	DL	RL	Accuracy Control	Precision Control
Aluminum	<20 ug/L	20	20		
Aluminum (RPD)	25.0 %				0 - 20
Antimony (RPD)	0 %				0 - 20
Antimony	<10 ug/L	5	10		
Arsenic	1.73 ug/L	0.5	1		
Arsenic (RPD)	1.6 %				0 - 20
Barium (RPD)	3.0 %				0 - 20
Barium	18.5 ug/L	2	10		
Beryllium	<0.5 ug/L	0.1	0.5		
Beryllium (RPD)	0 %				0 - 20
Boron (RPD)	3.5 %				0 - 20
Boron	<25 ug/L	5	25		
Cadmium	<1 ug/L	0.3	1		
Cadmium (RPD)	133 %				0 - 20
Calcium (RPD)	4.1 %				0 - 20
Calcium	4.98 mg/L	0.04	0.04		
Chromium	<1 ug/L	0.3	1		
Chromium (RPD)	0 %				0 - 20
Cobalt (RPD)	160 %				0 - 20
Cobalt	<1 ug/L	0.5	1		
Copper	<1 ug/L	0.5	1		
Copper (RPD)	28.0 %				0 - 20
Iron (RPD)	5.2 %				0 - 20
Iron	393 ug/L	10	20		
Lead	<1 ug/L	0.1	1		
Lead (RPD)	8.3 %				0 - 20
Magnesium (RPD)	1.7 %				0 - 20
Magnesium	2.28 mg/L	0.1	0.1		
Manganese	350 ug/L	0.2	1		
Manganese (RPD)	8.1 %				0 - 20
Nickel (RPD)	11 %				0 - 20
Nickel	<2.5 ug/L	0.5	2.5		
Potassium	2.22 mg/L	0.05	1		

Arkansas Department of Environmental Quality  
 5301 Northshore Drive  
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
 Ruehr@adeq.state.ar.us  
 501-682-0955

Potassium (RPD)	3.9 %			0 - 20
Selenium (RPD)	35.5 %			0 - 20
Selenium	<2 ug/L	0.5	2	
Silver	<5 ug/L	1	5	
Silver (RPD)	0 %			0 - 20
Sodium	3.55 mg/L	0.02	0.04	
Sodium (RPD)	0.1 %			0 - 20
Thallium	<2.5 ug/L	0.05	2.5	
Thallium (RPD)	0 %			0 - 20
Vanadium (RPD)	21.6 %			0 - 20
Vanadium	<2.5 ug/L	1	2.5	
Zinc	<3 ug/L	2	3	
Zinc (RPD)	20.2 %			0 - 20
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Aug 11 2015 12:53AM			

## Analytical Quality Control Results Report

<b>Batch: 15080703</b>	<b>ICP Metals - water (Diss.)</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*ICP Metals - water (Dissolved) DUP*

*Run: 1*

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Aluminum	<20 ug/L	20	20		
Aluminum (RPD)	73.1 %				0 - 20
Antimony (RPD)	0 %				0 - 20
Antimony	<5 ug/L	1	5		
Arsenic	1.10 ug/L	0.2	0.5		
Arsenic (RPD)	4.4 %				0 - 20
Barium (RPD)	0 %				0 - 20
Barium	8.47 ug/L	0.4	2		
Beryllium	<0.1 ug/L	0.04	0.1		
Beryllium (RPD)	0 %				0 - 20
Boron (RPD)	0.3 %				0 - 20
Boron	13.8 ug/L	2	5		
Cadmium	<0.1 ug/L	0.05	0.1		
Cadmium (RPD)	0 %				0 - 20
Calcium (RPD)	0.2 %				0 - 20
Calcium	5.04 mg/L	0.03	0.03		
Chromium	<0.5 ug/L	0.05	0.5		
Chromium (RPD)	0 %				0 - 20
Cobalt (RPD)	0 %				0 - 20
Cobalt	<0.5 ug/L	0.05	0.5		
Copper	<0.5 ug/L	0.2	0.5		
Copper (RPD)	182 %				0 - 20
Iron (RPD)	1.3 %				0 - 20
Iron	28.7 ug/L	5	20		
Lead	<0.3 ug/L	0.02	0.3		
Lead (RPD)	0 %				0 - 20
Magnesium (RPD)	1.8 %				0 - 20
Magnesium	2.32 mg/L	0.01	0.02		
Manganese	19 ug/L	0.07	0.3		
Manganese (RPD)	2.3 %				0 - 20
Nickel (RPD)	20 %				0 - 20
Nickel	<0.5 ug/L	0.15	0.5		
Potassium	2.26 mg/L	0.01	0.02		



Arkansas Department of Environmental Quality  
 5301 Northshore Drive  
 North Little Rock, AR 72118

Laboratory Contact: Jeff Ruehr  
 Ruehr@adeq.state.ar.us  
 501-682-0955

Potassium (RPD)	0.2 %			0 - 20
Selenium (RPD)	3.8 %			0 - 20
Selenium	<1 ug/L	0.2	1	
Silicon Dioxide	2.95 mg/L	0.01	0.05	
Silicon Dioxide (RPD)	0.1 %			0 - 20
Silver (RPD)	0 %			0 - 20
Silver	<0.5 ug/L	0.02	0.5	
Sodium	3.02 mg/L	0.01	0.02	
Sodium (RPD)	3.9 %			0 - 20
Thallium (RPD)	0 %			0 - 20
Thallium	<0.5 ug/L	0	0.5	
Vanadium (RPD)	5.1 %			0 - 20
Vanadium	<0.5 ug/L	0.3	0.5	
Zinc (RPD)	0 %			0 - 20
Zinc	<1 ug/L	0.3	1	
Hardness	22 mg/L	1	1	
Hardness (RPD)	0 %			0 - 20
Dilution Factor	1			
Analyzed By	Robert Graddy			
Analysis Date/Time	Aug 7 2015 3:58AM			

## Analytical Quality Control Results Report

<b>Batch: 15062907</b>	<b>Turbidity - water</b>
<b>WS-001</b>	<b>LIMS ID: 2015-1648</b>

*Turbidity - water DUP*

*Run: 1*

<i>Parameter</i>	<i>Result</i>	<i>DL</i>	<i>RL</i>	<i>Accuracy Control</i>	<i>Precision Control</i>
Turbidity	8.65 NTU	0.02	0.02		
Turbidity (RPD)	0.5 %				0 - 20
Dilution Factor	1				
Analyzed By	Kathryn Hattenhauer				
Analysis Date/Time	6/29/2015 11:24				