

PH CALIBRATION - 0925 7/27/2016

PRECALIBRATION

7.0 - 6.93
4.0 - 3.96

YSI pH100 METER

MODEL No.: pH100
SERIAL No.: JC04242
LOT CODE: 10J1

CALIBRATION

7.0 - 6.98
4.0 - 4.02
EFF. = 95.8

POST CALIBRATION

7.0 = 6.87 @ 1122
4.0 = 3.92

ARRIVAL: 09:20

DEPARTURE: 11:45

PH BUFFERS

4.00 - Lot # 5GH467, EXPIRES: 8/2017
7.00 - Lot # 6GE010, EXPIRES: 9/2018
10.00 - Lot # 5GF949, EXPIRES: 6/2017